

Supplementary Figure S1: Receiver operating characteristic (ROC) curve analysis of the Lumipulse® G SARS-CoV-2 Ag. The automated antigen detection test achieved an area under the ROC curve (AUC) value of 0.893 ± 0.021 compared to the Abbott *m2000* SARS-CoV-2 PCR detection as Gold Standard.

