



OPEN ACCESS

EDITED AND REVIEWED BY
Gianni Ciofani,
Italian Institute of Technology (IIT), Italy

*CORRESPONDENCE
Juntao Liu,
✉ liujuntao@mail.ie.ac.cn

RECEIVED 16 December 2024
ACCEPTED 19 December 2024
PUBLISHED 06 January 2025

CITATION
Liu J, Yang Z, Wang Y, Wang L and Li Z (2025)
Editorial: Micro/nano devices and technologies
for neural science and medical applications.
Front. Bioeng. Biotechnol. 12:1545853.
doi: 10.3389/fbioe.2024.1545853

COPYRIGHT
© 2025 Liu, Yang, Wang, Wang and Li. This is an
open-access article distributed under the terms
of the [Creative Commons Attribution License
\(CC BY\)](#). The use, distribution or reproduction in
other forums is permitted, provided the original
author(s) and the copyright owner(s) are
credited and that the original publication in this
journal is cited, in accordance with accepted
academic practice. No use, distribution or
reproduction is permitted which does not
comply with these terms.

Editorial: Micro/nano devices and technologies for neural science and medical applications

Juntao Liu^{1*}, Zhugen Yang², Yang Wang³, Li Wang⁴ and Ziyue Li⁵

¹Aerospace Information Research Institute, Chinese Academy of Sciences, Beijing, China, ²Centre for Water, Environment and Development, Cranfield University, Cranfield, Oxfordshire, United Kingdom, ³School of Biological Science and Medical Engineering, Beihang University, Beijing, China, ⁴Faculty of Mechanical Engineering, Qilu University of Technology, Ji'nan, China, ⁵University of Connecticut Health Center, University of Connecticut, Storrs, CT, United States

KEYWORDS

micro/nano devices, neural science, brain computer interface, nanotechnology, life science

Editorial on the Research Topic

[Micro/nano devices and technologies for neural science and medical applications](#)

Research on micro/nano devices and technologies represents a significant Frontier at the intersection of information science and life sciences, holding substantial strategic importance and promising application prospects in the fields of neural science and medical applications (Liu et al., 2020). With the rapid advancement of micro/nano processing technology, innovative intelligent, miniaturized, and integrated devices are emerging. These devices offer distinct advantages in detection and regulation. Notably, integrating micro/nano devices with neural science and clinical medicine can address scientific frontiers while fostering new research hotspots.

Epilepsy is a major neurological disorder that affects over sixty million individuals worldwide, severely impacting their health and quality of life (Bernhardt et al., 2019). To elucidate the mechanisms underlying epilepsy as well as its treatment options, it is essential to investigate changes in neural activity within relevant neural circuits. Implantable microelectrode arrays—capable of high-quality signal recording and neural information decoding—exhibit considerable potential for application in brain-computer interfaces (Wang et al., 2024). Han et al. designed and fabricated an implantable microelectrode array specifically for electrophysiological signal detection and analysis within the striatal region of the basal ganglia in epileptic rats. The analysis of electrophysiological data from the striatum during episodes of epilepsy provides valuable insights into the dynamic processes governing striatal neural activity during both initial onset phases as well as latent periods associated with temporal lobe epilepsy. This understanding contributes to unraveling the neural mechanisms involved in epilepsy while promoting advancements in related therapeutic approaches for this condition.

Pain is an emotional and unpleasant sensory experience that can lead to significant physiological and psychological effects across various aspects of life and work. Recent advancements in nanotechnology have paved the way for innovative pain relief strategies utilizing diverse nanomaterials and targeted surface

modifications (Shi et al., 2023). Zhu et al. summarized the status of the research and global trends of nanotechnology in relation to pain management. Neuromodulation represents another therapeutic approach for alleviating pain by modulating abnormal nerve activity through light, sound, or electrical stimulation. Nanomaterials such as carbon nanotubes, graphene, piezoelectric ceramics, and magnetic metals have garnered extensive research interest for use in implantable or wearable devices designed for neuromodulation. In conclusion, emerging novel drugs and technologies present a wide array of options for effective pain management.

The detection of biomarkers holds significant importance for disease diagnosis and treatment as well as life science research (Sun et al., 2022). Biosensors are specialized sensors capable of detecting a broad spectrum of biological and chemical substances. The integration of nanotechnology into the field of biosensors has enhanced their sensitivity along with other critical properties, leading to the development of innovative biosensing solutions. Focusing on E2 specifically, Wang et al. reviewed the extensive applications of new functional nanomaterials within various biosensors while discussing whether modified bioreceptors or energy exchange interfaces enable these functional nanomaterials to enhance sensing performance through improved signal conversion rates, signal amplification capabilities, and biocompatibility.

Nanoenzymes represent a novel class of nanomaterials exhibiting enzyme-like activity that has attracted considerable attention within the biomedical field in recent years. Compared to traditional biological enzymes, nanoenzymes demonstrate superior stability alongside greater heat resistance and broader pH adaptability ranges; these characteristics underscore their substantial potential in various applications (Christodoulou, et al., 2023). Chen et al. classified metal-organic framework (MOF)-based nanozymes, analyzed their application in stroke treatment and highlighted the challenges in developing efficient antioxidant nanozyme systems for stroke treatment.

Micro-nano technology is currently at the forefront of technological advancement, significantly enhancing detection and regulation capabilities. We can leverage micro-nano technology to develop new devices and equipment tailored to specific needs. To address the demand for deep brain detection and regulation, it is essential to create miniaturized, high-throughput implantable brain-computer interfaces (Yang et al., 2024). In response to point-of-care testing (POCT) requirements, we must focus on developing biosensors and devices that offer rapid detection times, high sensitivity, and user-friendly operation. Utilizing micro-nano technology allows for the design and development of innovative miniaturized sensor devices capable of achieving previously unattainable levels of detection. For instance, the nervous system transmits information through two fundamental modes: electrical signals from nerve cells and

chemical transmitters (dual-mode). The dual-mode neural microelectrode array can mitigate information loss associated with single-mode recording. This advancement holds significant scientific importance and application value in researching neural diseases and advancing our understanding of brain science. We believe that through ongoing development and innovation, these technologies will play a pivotal role in driving progress within neuroscience and clinical medicine.

Author contributions

JL: Writing–original draft. ZY: Writing–review and editing. YW: Writing–review and editing. LW: Writing–review and editing. ZL: Writing–review and editing.

Funding

The author(s) declare that financial support was received for the research, authorship, and/or publication of this article. This work was sponsored by STI 2030 – Major Projects 2021ZD0201603, the Joint Foundation Programme of the Chinese Academy of Sciences (8091A170201), National Natural Science Foundation of China (No. T2293730, T2293731, T2293734, 62,471,291, 62,121,003, 61,960,206,012, 62,333,020 and 62,171,434), the National Key Research and Development Program of China (No. 2022YFC2402501, 2022YFB3205602).

Conflict of interest

The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

Generative AI statement

The author(s) declare that no Generative AI was used in the creation of this manuscript.

Publisher's note

All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.

References

- Bernhardt, B. C., Fadaie, F., Liu, M., Caldairou, B., Gu, S., Jefferies, E., et al. (2019). Temporal lobe epilepsy: hippocampal pathology modulates connectome topology and controllability. *Neurology*, 92 (19). doi:10.1212/WNL.0000000000007447
- Christodoulou, I., Lyu, P., Soares, C. V., Patriarche, G., Maurin, G., Gref, R., et al. (2023). Nanoscale iron-based metal-organic frameworks: incorporation of functionalized drugs and degradation in biological media. *Int. J. Mol. Sci.* 24 (4), 3362. doi:10.3390/ijms24043362
- Liu, S. J., Zhao, Y., Hao, W. T., Zhang, X. D., and Ming, D. (2020). Micro and nanotechnology for neural electrode-tissue interfaces. *Biosens. Bioelectron.* 170, doi:10.1016/j.bios.2020.112645
- Shi, S. Y., Ren, H. Z., Xie, Y. J., Yu, M., Chen, Y., and Yang, L. (2023) Engineering advanced nanomedicines against central nervous system diseases. *Mater. Today*, 69: 355–392. doi:10.1016/j.mattod.2023.08.005
- Sun, S., Luo, J. P., Zhu, Y. X., Kong, F., Mao, G., Ming, T., et al. (2022). Multifunctional self-driven origami paper-based integrated microfluidic chip to detect CRP and PAB in whole blood. *Bios. Bioelectron.* 208: 114225. doi:10.1016/j.bios.2022.114225
- Wang, Y., Han, M. Q., Xu, Z. J., Lv, S., Yang, G., Mo, E., et al. (2024). Chronic implantable flexible serpentine probe reveals impaired spatial coding of place cells in epilepsy. *Natl. Sci. Rev.*, nwae402. doi:10.1093/nsr/nwae402
- Yang, G. C., Wang, Y. D., Mo, F., Xu, Z., Lu, B., Fan, P., et al. (2024). PEDOT-Citrate/SIKVAV modified bioaffinity microelectrode arrays for detecting theta rhythm cells in the retrosplenial cortex of rats under sensory conflict. *Sensor. Actuat. B-Chem.* 399, 134802. doi:10.1016/j.snb.2023.134802