Check for updates

OPEN ACCESS

EDITED BY Janardhan Reddy Koduru, Kwangwoon University, Republic of Korea

REVIEWED BY Chang-Feng Chi, Zhejiang Ocean University, China Roman Shafigulin, Samara University, Russia Rama Rao Karri, University of Technology Brunei, Brunei

*CORRESPONDENCE Yan Li, ⊠ liyanxjtu@xjtu.edu.cn Jianli Liu, ⊠ jiliu@nwu.edu.cn

SPECIALTY SECTION This article was submitted to Analytical Chemistry, a section of the journal Frontiers in Chemistry

RECEIVED 22 February 2023 ACCEPTED 29 March 2023 PUBLISHED 18 April 2023

CITATION

Peng H, Yang X, Fang H, Zhang Z, Zhao J, Zhao T, Liu J and Li Y (2023), Simultaneous effect of different chromatographic conditions on the chromatographic retention of pentapeptide derivatives (HGRFG and NPNPT). *Front. Chem.* 11:1171824. doi: 10.3389/fchem.2023.1171824

COPYRIGHT

© 2023 Peng, Yang, Fang, Zhang, Zhao, Zhao, Liu and Li. This is an open-access article distributed under the terms of the **Creative Commons Attribution License** (CC BY). The use, distribution or reproduction in other forums is permitted, provided the original author(s) and the copyright owner(s) are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.

Simultaneous effect of different chromatographic conditions on the chromatographic retention of pentapeptide derivatives (HGRFG and NPNPT)

Huan Peng^{1,2}, Xiangrong Yang^{2,3}, Huanle Fang⁴, Zhongqi Zhang⁵, Jinli Zhao⁵, Te Zhao⁶, Jianli Liu^{2,4}* and Yan Li¹*

¹Center for Brain Science, The First Affiliated Hospital of Xi' an Jiaotong University, Xi'an, Shaanxi, China, ²College of Life Science, Northwest University, Xi'an, Shaanxi, China, ³Kangya of Ningxia Pharmaceutical Co., Ltd., Yinchuan, China, ⁴Medical College, Peihua University, Xi'an, Shaanxi, China, ⁵Department of Polypeptide Engineering, Active Protein and Polypeptide Engineering Center of Xi'an Hui Kang, Xi'an, Shaanxi, China, ⁶College of Electronic Engineering, Xidian University, Xi'an, Shaanxi, China

Introduction: Oligopeptides exhibit great prospects for clinical application and its separation is of great importance in new drug development.

Methods: To accurately predict the retention of pentapeptides with analogous structures in chromatography, the retention times of 57 pentapeptide derivatives in seven buffers at three temperatures and four mobile phase compositions were measured via reversed-phase high-performance liquid chromatography. The parameters (k_{HA} , k_A , and pKa) of the acid-base equilibrium were obtained by fitting the data corresponding to a sigmoidal function. We then studied the dependence of these parameters on the temperature (T), organic modifier composition (φ , methanol volume fraction), and polarity (P_m^N parameter). Finally, we proposed two six-parameter models with (1) pH and T and (2) pH and φ or P_m^N as the independent variables. These models were validated for their prediction capacities by linearly fitting the predicted retention factor k-value and the experimental k-value.

Results: The results showed that $logk_{HA}$ and $logk_A$ exhibited linear relationships with 1/T, φ or P_m^N for all pentapeptides, especially for the acid pentapeptides. In the model of pH and T, the correlation coefficient (R²) of the acid pentapeptides was 0.8603, suggesting a certain prediction capability of chromatographic retention. Moreover, in the model of pH and φ or P_m^N , the R² values of the acid and neutral pentapeptides were greater than 0.93, and the average root mean squared error was approximately 0.3, indicating that the k-values could be effectively predicted.

Abbreviations: ADME, absorption, distribution, metabolism, and excretion; RP-HPLC, reversed-phase high-performance liquid chromatography; RPLC, reversed-phase liquid chromatography; SPPS, solid-phase synthesis; TFA, trifluoroacetic acid; pKa, acid dissociation constant; T, temperature; φ , organic modifier volume fraction; ${}^{W}_{W} P H^{W}$, the pH measured in the aqueous buffer; ${}^{S}_{W} p H^{W}$, the pH measured in the aqueous buffer; ${}^{S}_{W} p H^{W}$, the pH measured in the methanol–water mixture; ${}^{S}_{S} p H^{S}$, the pH relative to the organic modifier–water solvent; $\log k_{W}$, extrapolation retention factor; k, retention factor; k_{HA} , the limiting retention factor of the protonated form; k_A , the limiting retention factor; k, retention factor; k_{HA} , the D-value of enthalpy; ΔS^0 , the D-value of entropy; Φ , the phase ratio; P_m^N , the standard polarity parameter of the mobile phase; P_s^N , the standard polarity parameter of the stationary phase; ($log k_0$, the retention factor when the polarity of the mobile phase is the same as that of the stationary phase; X, the variable describing the change in the mobile phase; R^2 , the correlation coefficient; and RMSE, the root mean squared error.

Discussion: In summary, the two six-parameter models were appropriate to characterize the chromatographic retention of amphoteric compounds, especially the acid or neutral pentapeptides, and could predict the chromatographic retention of pentapeptide compounds.

KEYWORDS

chromatographic retention, pentapeptides, six-parameter model, retention factor, prediction capacity



1 Introduction

Active peptides are common small-molecule compounds in nature and generally possess invaluable medicinal value (Abdelhedi and Nasri, 2019; Suo et al., 2022). Peptides have a wide range of bioactivities and can be divided into two categories according to different sources: (1) endogenous peptides from precursor proteins and secreted cells and (2) exogenous peptides from enzymatic hydrolysis or synthesis (Wang et al., 2022). The oligopeptides produced by the enzymatic hydrolysis of animal proteins have been reported to exhibit outstanding hypotensive effects by inhibiting the angiotensin-I converting enzyme (Abdelhedi et al., 2017; Qiao et al., 2022). Moreover, the oligopeptides extracted from tea and brewer's spent grain had excellent hypolipidemic activities (Ferreira et al., 2022; Ye et al., 2023), and the oligopeptides isolated from Siberian sturgeon cartilage could treat chronic diseases caused by oxidative stress (Sheng et al., 2022). More importantly, oligopeptides exhibit great prospects for clinical application due to their high degree of affinity and specificity and easy absorption (Zhang et al., 2020; Sitkov et al., 2021).

Oligopeptide separations by reversed-phase high-performance liquid chromatography (RP-HPLC) are extremely common (Ofosu et al., 2021; Samtiya et al., 2021; Waili et al., 2021), and the chromatographic retention of oligopeptides in RP-HPLC is driven by hydrophobic interactions (Sousa et al., 2021). Combining the molecular structure of compounds with the parameters describing the properties of chromatographic mobile and stationary phases, functional relationships could be obtained (Nie et al., 2022). These relationships can be used to analyze and predict the chromatographic behavior of other compounds (Janicka et al., 2020; Nie et al., 2022) and evaluate the pharmacokinetics and biochemical properties of drugs, such as absorption, distribution, metabolism, and excretion (ADME) *in vivo* (Langyan et al., 2021). It can also preliminarily determine the solubility, lipophilicity, bioaccumulation, and toxicity of compounds *in vivo*, which is of great significance in the field of new drug molecule development, especially in the analysis of the chemical properties of peptides *in vivo*.

The acid dissociation constant (pKa) is an elementary parameter in the analysis of drugs and strongly affects their pharmacokinetics and biochemical properties by characterizing the degree of ionization of drug molecules in solution at different pH values (Besleaga et al., 2021). pKa determines the existing form of compounds in the medium and their solubility, lipophilicity, permeability, bioaccumulation, and toxicity (Konçe et al., 2019; Xie et al., 2022); these characteristics play a particularly important role in the drug development process (Bergazin et al., 2021). Accurate prediction of the pKa value of organic compounds is highly important in numerous fields, especially in the development of new drugs (Xiong et al., 2022; Zhang et al., 2022). However, accurate prediction of the pKa for drug-like molecules is also a tremendous challenge in chemistry (Zhang et al., 2022).

Due to its high-resolution ratio, selectivity, and reproducibility, RP-HPLC is the most extensive and central technique in the analysis and separation of a wide range of compounds and the study of the pKa values of drug molecules (D'Archivio, 2019; Yılmaz Ortak and Cubuk Demiralay, 2019). Apart from molecular structure, numerous factors in chromatographic analysis programs have an important influence on retention time, such as the pH of the mobile phase, column temperature, mobile phase composition, and type of chromatographic column (Huang et al., 2019; Tsui et al., 2019; Annadi et al., 2022; Shi et al., 2022). The chromatographic conditions can be adjusted and

Aqueous buffer		TABLE 1 $_{W}^{S}$ pH ^W and $_{S}^{S}$ pH ^S in the mobile phase of high-performance liquid chromatography. $_{W}^{S}$ pH ^W													
Aqueous buffer			-02	Whit		2					-02				
		2:	5°C			3.	5°C			4	5°C				
	8%	10%	12%	14%	8%	10%	12%	14%	8%	10%	12%	14%			
13.46 mM TFA	1.73	1.75	1.75	1.76	1.75	1.77	1.77	1.78	1.87	1.89	1.89	1.89			
1.08 mM H ₃ Cit+0.04 mM Na ₃ Cit	3.07	3.09	3.10	3.15	3.09	3.12	3.13	3.14	3.12	3.16	3.18	3.19			
1.92 mM H ₃ Cit+0.31 mM Na ₃ Cit	4.02	4.03	4.07	4.11	4.02	4.06	4.08	4.10	4.07	4.09	4.12	4.13			
0.74 mM H ₃ Cit+0.26 mM Na ₃ Cit	5.07	5.09	5.12	5.12	5.05	5.09	5.13	5.14	5.05	5.11	5.14	5.19			
0.55 mM Na ₂ HPO ₄ +12.51 mM NaH ₂ PO ₄	6.04	6.08	6.15	6.18	6.03	6.08	6.13	6.15	6.04	6.08	6.13	6.17			
1.36 mM Na ₂ HPO ₄ +2.63 mM NaH ₂ PO ₄	7.09	7.12	7.18	7.24	7.08	7.10	7.15	7.17	7.02	7.09	7.15	7.18			
2.11 mM Na ₂ HPO ₄ +1.02 mM NaH ₂ PO ₄	8.08	8.15	8.17	8.20	8.04	8.10	8.11	8.21	8.00	8.03	8.07	8.12			
Aqueous buffer		2	5°C			3	5°C			4	5°C				
	20%	22%	24%	26%	20%	22%	24%	26%	20%	22%	24%	26%			
13.46 mM TFA	1.76	1.76	1.77	1.79	1.79	1.80	1.81	1.84	1.91	1.94	1.97	1.98			
1.08 mM H ₃ Cit+0.04 mM Na ₃ Cit	3.17	3.18	3.21	3.23	3.21	3.21	3.23	3.23	3.22	3.24	3.27	3.29			
1.92 mM H ₃ Cit+0.31 mM Na ₃ Cit	4.17	4.19	4.21	4.25	4.17	4.22	4.25	4.30	4.21	4.24	4.26	4.31			
0.74 mM H ₃ Cit+0.26 mM Na ₃ Cit	5.25	5.27	5.31	5.35	5.22	5.26	5.32	5.36	5.27	5.33	5.37	5.41			
0.55 mM Na ₂ HPO ₄ +12.51 mM NaH ₂ PO ₄	6.27	6.31	6.38	6.41	6.24	6.29	6.34	6.38	6.27	6.33	6.37	6.42			
1.36 mM Na ₂ HPO ₄ +2.63 mM NaH ₂ PO ₄	7.32	7.37	7.41	7.46	7.29	7.33	7.40	7.41	7.30	7.34	7.40	7.45			
2.11 mM Na ₂ HPO ₄ +1.02 mM NaH ₂ PO ₄	8.33	8.40	8.43	8.45	8.30	8.33	8.38	8.45	8.25	8.31	8.35	8.36			
				sspHs											
Aqueous buffer		2	5°C			3.	5°C			4	5°C				
	8%	10%	12%	14%	8%	10%	12%	14%	8%	10%	12%	14%			
13.46 mM TFA												1.87			
												3.17			
												4.11			
												5.17			
	6.03	6.07	6.14	6.16	6.02	6.07	6.12	6.13	6.03	6.07	6.12	6.15			
1.36 mM Na ₂ HPO ₄ +2.63 mM NaH ₂ PO ₄	7.08	7.11	7.17	7.22	7.07	7.09	7.14	7.15	7.01	7.08	7.14	7.16			
2.11 mM Na ₂ HPO ₄ +1.02 mM NaH ₂ PO ₄	8.07	8.14	8.16	8.18	8.03	8.09	8.10	8.19	7.99	8.02	8.06	8.10			
Aqueous buffer															
-	20%	22%	24%	26%	20%	22%	24%	26%	20%	22%	24%	26%			
13.46 mM TFA	1.73	1.73	1.73	1.75	1.76	1.77	1.77	1.80	1.88	1.91	1.93	1.94			
1.08 mM H ₃ Cit+0.04 mM Na ₃ Cit	3.14	3.15	3.17	3.19	3.18	3.18	3.19	3.19	3.19	3.21	3.23	3.25			
		4.16	4.17	4.21	4.14	4.19	4.21	4.26	4.18	4.21	4.22	4.27			
1.92 mM H ₃ Cit+0.31 mM Na ₃ Cit	4.14														
1.92 mM H ₃ Cit+0.31 mM Na ₃ Cit 0.74 mM H ₃ Cit+0.26 mM Na ₃ Cit	5.22	5.24	5.27	5.31	5.19	5.23	5.28	5.32	5.24	5.30	5.33	5.37			
			5.27 6.34	5.31 6.37	5.19 6.21	5.23 6.26	5.28 6.30	5.32 6.34	5.24 6.24	5.30 6.30	5.33 6.33				
0.74 mM H ₃ Cit+0.26 mM Na ₃ Cit	5.22	5.24										5.37			
	1.08 mM H₃Cit+0.4 mM Na₃Cit 1.92 mM H₃Cit+0.31 mM Na₃Cit 0.74 mM H₃Cit+0.26 mM Na₃Cit 1.35 mM Na₂HPO₄+1.251 mM NaH₂PO₄ 2.11 mM Na₂HPO₄+1.02 mM NaH₂PO₄ Aqueous buffer 1.36 mM H₃Cit+0.4 mM Na₃Cit 0.74 mM H₃Cit+0.4 mM Na₃Cit 1.08 mM H₃Cit+0.4 mM Na₃Cit 0.74 mM H₃Cit+0.26 mM Na₃Cit 0.74 mM H₃Cit+0.26 mM Na₄Cit 1.36 mM Na₂HPO₄+1.251 mM NaH₂PO₄ 1.36 mM Na₂HPO₄+1.02 mM NaH₂PO₄ 1.36 mM Na₂HPO₄+1.02 mM NaH₂PO₄ 1.31 mM Na₂HPO₄+1.02 mM NaH₂PO₄ 1.36 mM Na₂HPO₄+1.02 mM NaH₂PO₄ 1.36 mM Na₂HPO₄+1.02 mM NaH₂PO₄ 1.36 mM H₃Cit+0.31 mM Na₃Cit 1.92 mM H₃Cit+0.31 mM Na₃Cit 1.92 mM H₃Cit+0.31 mM Na₃Cit 1.92 mM H₃Cit+0.26 mM Na₃Cit 1.92 mM H₃Cit+0.31 mM Na₃Cit 0.74 mM H₃Cit+0.26 mM Na₃Cit 0.74 mM H₃Cit+0.26 mM Na₃Cit 1.92 mM H₃Cit+0.26 mM Na₄Cit 0.74 mM H₃Cit+0.26 mM Na₄Cit 0.74 mM H₃Cit+0.26 mM Na₄Cit 1.36 mM Na₂HPO₄+1.251 mM Na₄Cit 1.36 mM Na₂HPO₄+1.20 mM NaH₂PO₄ 1.36 mM Na₂HPO₄+1.02 mM NaH₂PO₄ 1.36 mM Na₂HPO₄+1.02 mM NaH₂PO₄	1.08 mM H3Cit+0.04 mM Na3Cit3.071.92 mM H3Cit+0.31 mM Na3Cit4.020.74 mM H3Cit+0.26 mM Na3Cit5.070.55 mM Na2HPO4+1.251 mM NaH2PO46.041.36 mM Na2HPO4+1.02 mM NaH2PO48.08Aqueous buffer11.08 mM H3Cit+0.21 mM Na4L2PO41.071.09 mJ Aight1.071.01 mJ Na2HPO4+1.02 mM Na4L2PO41.071.02 mJ H3Cit+0.04 mJ Na3Cit4.171.02 mJ H3Cit+0.25 mJ Na4L2PO45.250.55 mJ Na2HPO4+1.251 mJ Na4L2PO46.271.36 mJ Na2HPO4+1.251 mJ Na4L2PO46.271.36 mJ Na2HPO4+1.02 mJ Na4L2PO47.322.11 mJ Na2HPO4+1.02 mJ Na4L2PO48.331.36 mJ H3Cit+0.41 mJ Na4Cit3.061.192 mJ H3Cit+0.41 mJ Na4Cit3.061.192 mJ H3Cit+0.26 mJ Na4L2PO43.061.36 mJ Na2HPO4+1.02 mJ Na4L2PO43.061.108 mJ H3Cit+0.26 mJ Na3Cit3.061.192 mJ H3Cit+0.31 mJ Na3Cit3.061.92 mJ H3Cit+0.31 mJ Na4Cit3.061.92 mJ H3Cit+0.26 mJ Na4Cit3.061.92 mJ H3Cit+0.26 mJ Na4Cit5.061.92 mJ H3Cit+0.26 mJ Na4Cit5.061.93 mJ Na2HPO4+1.02 mJ Na4Cit5.061.94 mJ Na2HPO4+1.02 mJ Na4Cit5.061.95 mJ N	Indextrem Instruct Instruct Instruct 1.08 mM H_3Cit+0.31 mM Na_3Cit 3.09 4.03 4.03 0.74 mM H_3Cit+0.26 mM Na_3Cit 5.09 5.09 5.09 0.55 mM Na_2HPO_4+12.51 mM NaH_2PO4 6.04 6.08 6.03 1.36 mM Na_2HPO_4+1.02 mM NaH_2PO4 8.08 8.15 6.04 2.11 mM Na_2HPO_4+1.02 mM NaH_2PO4 8.08 8.15 6.04 Aqueous buffer 20% 2.29% 7.12 1.08 mM H_3Cit+0.04 mM Na_3Cit 3.17 3.18 7.12 1.08 mM H_3Cit+0.31 mM Na_3Cit 4.17 4.19 7.37 0.74 mM H_3Cit+0.26 mM Na_3Cit 6.27 6.31 7.37 1.36 mM Na_2HPO4+1.251 mM NaH_2PO4 7.32 7.37 7.37 2.11 mM Na_2HPO4+1.02 mM NaH_2PO4 8.33 8.40 7.32 1.36 mM TFA 1.72 1.74 1.74 1.36 mM Na_2HPO4+1.02 mM NaH_2PO4 8.33 8.40 1.92 mM H_3Cit+0.31 mM Na_3Cit 3.06 3.08 1.92 mM H_3Cit+0.31 mM Na_3Cit 3.06 5.08	1.08 mM H3Cit+0.04 mM Na3Cit3.073.093.101.92 mM H3Cit+0.31 mM Na3Cit4.024.034.070.74 mM H3Cit+0.26 mM Na4Cit5.075.095.120.55 mM Na2HPQ4+1.251 mM NaH2PQ46.046.086.151.36 mM Na2HPQ4+2.63 mM NaH2PQ48.088.158.172.11 mM Na2HPQ4+1.02 mM NaH2PQ48.088.158.17Aqueous buffer20%22%24%1.08 mM H3Cit+0.04 mM Na3Cit3.173.183.211.92 mM H3Cit+0.31 mM Na4Cit4.174.194.210.74 mM H3Cit+0.26 mM Na4Cit5.255.275.310.55 mM Na2HPQ4+12.51 mM NaH2PQ46.276.316.381.36 mM Na2HPQ4+12.51 mM NaH2PQ47.327.347.412.11 mM Na2HPQ4+1.02 mM NaH2PQ48.338.408.331.36 mM Na2HPQ4+1.02 mM NaH2PQ48.338.408.331.36 mM Na2HPQ4+1.02 mM NaH2PQ48.338.408.311.10 Ma2HPQ4+1.02 mM Na42PQ41.721.741.741.08 mM H3Cit+0.31 mM Na3Cit3.063.083.091.92 mM H3Cit+0.26 mM Na3Cit3.063.083.091.92 mM H3Cit+0.26 mM Na3Cit3.065.013.161.92 mM H3Cit+0.27 mM Na42PQ43.023.083.091.92 mM H3Cit+0.26 mM Na3Cit3.065.013.161.92 mM H3Cit+0.26 mM Na42PQ44.024.024.021.92 mM H3Cit+0.26 mM Na42PQ45.065.085.111.92 mM H3Cit+0.26 mM Na42PQ46.036.07 <t< td=""><td>Intermediate Interme</br></br></br></br></br></br></br></br></br></br></br></td><td>Image: region of the state o</td><td>1.08 mM H_3Cit+0.04 mM Na_3Cit3.073.093.103.113.073.121.92 mM H_3Cit+0.31 mM Na_3Cit4.024.034.074.114.024.060.74 mM H_3Cit+0.26 mM Na_4Cit5.075.095.125.125.035.030.55 mM Na_4PPO_4+1.25 1 mM NAH_PO6.046.086.156.186.036.101.16 mM Na_4PPO_4+0.26 mM NAH_PO7.097.127.187.247.087.102.11 mM Na_4PO_4+1.02 mM NAH_PO8.088.158.178.208.048.101.16 mM TA1.761.761.771.791.791.791.791.08 mM H_3Cit+0.04 mM Na_5Cit3.173.183.213.213.213.211.92 mM H_3Cit+0.31 mM Na_Cit4.174.194.214.254.174.220.74 mM H_3Cit+0.26 mM Na_4Cit5.255.275.315.355.235.245.241.92 mM H_3Cit+0.26 mM Na_4Cit7.337.337.447.457.332.11 mM Na_4PO_4+1.25 mM NAH_PO6.236.316.388.408.338.403.55 mM Na_4PO_4+1.25 mM NAH_PO7.337.347.447.457.341.16 mM Na_4PO_4+1.25 mM NAH_PO7.337.447.447.457.342.11 mM Na_4PO_4+1.25 mM NAH_PO7.338.408.438.458.438.453.16 mM Na_4PO_4+1.25 mM NAH_PO7.347.447.447.447.451.26 mM H_3Cit+0.41 mM Na_5Cit3.06<</td><td>1.08 mM H₂Cit+0.04 mM Na₃Cit3.073.093.103.113.093.123.131.92 mM H₂Cit+0.31 mM Na₃Cit5.075.095.125.125.055.095.130.55 mM Na₃HPO₄+1.25 1mM NaH₂PO6.046.086.156.186.036.013.131.36 mM Na₃HPO₄+1.25 1mM NaH₂PO7.097.127.187.247.087.017.152.11 mM Na₃HPO₄+1.02 mM NaH₂PO6.096.016.166.046.013.167.01Aqueous buffer20%2.2%2.4%6.046.044.024.024.021.08 m M₂Cit+0.41 mM Na₃Cit3.173.183.213.233.213.233.211.09 m M₂Cit+0.42 mM Na₃Cit3.173.183.213.233.213.233.213.231.09 m M₃Cit+0.41 mM Na₃Cit3.173.183.213.233.213.233.213.233.231.09 m M₃Cit+0.42 mM Na₃Cit3.173.183.213.233.233.233.233.233.231.09 m M₃Cit+0.42 mM Na₄Cit3.133.143.123.143.143.143.143.143.141.09 m M₃Cit+0.42 mM Na₄Cit3.133.143.143.143.143.143.143.141.01 m Na₄HPO₄+1.25 mM Na₄Dit3.143.143.143.143.143.143.141.10 m Na₄HPO₄+1.25 mM Na₄Dit3.143.143.143.14<</td><td>1.08 mM H,Cit+0.04 mM Na,Cit3.073.093.003.103.153.093.123.103.141.92 mM H,Cit+0.31 mM Na,Cit4.024.034.074.114.024.064.034.070.74 mM H,Cit+0.26 mM Na,Cit5.095.095.125.125.125.106.186.186.186.186.186.186.186.186.196.196.197.127.137.147.1</td><td>1.08 mM H,Cit+0.04 mM NayCit3.073.093.093.103.103.113.143.121.92 mM H,Cit+0.31 mM NayCit4.024.034.074.114.024.064.084.014.070.74 mM H,Cit+0.26 mM NayEit5.075.075.086.0</td><td>1.08 mM H,Girle 04 mM Na,Gir3.073.093.073.093.073.093.013.093.01</td><td>108 mM Highen Mangenia500</td></t<>	Intermediate 	Image: region of the state o	1.08 mM H_3Cit+0.04 mM Na_3Cit3.073.093.103.113.073.121.92 mM H_3Cit+0.31 mM Na_3Cit4.024.034.074.114.024.060.74 mM H_3Cit+0.26 mM Na_4Cit5.075.095.125.125.035.030.55 mM Na_4PPO_4+1.25 1 mM NAH_PO6.046.086.156.186.036.101.16 mM Na_4PPO_4+0.26 mM NAH_PO7.097.127.187.247.087.102.11 mM Na_4PO_4+1.02 mM NAH_PO8.088.158.178.208.048.101.16 mM TA1.761.761.771.791.791.791.791.08 mM H_3Cit+0.04 mM Na_5Cit3.173.183.213.213.213.211.92 mM H_3Cit+0.31 mM Na_Cit4.174.194.214.254.174.220.74 mM H_3Cit+0.26 mM Na_4Cit5.255.275.315.355.235.245.241.92 mM H_3Cit+0.26 mM Na_4Cit7.337.337.447.457.332.11 mM Na_4PO_4+1.25 mM NAH_PO6.236.316.388.408.338.403.55 mM Na_4PO_4+1.25 mM NAH_PO7.337.347.447.457.341.16 mM Na_4PO_4+1.25 mM NAH_PO7.337.447.447.457.342.11 mM Na_4PO_4+1.25 mM NAH_PO7.338.408.438.458.438.453.16 mM Na_4PO_4+1.25 mM NAH_PO7.347.447.447.447.451.26 mM H_3Cit+0.41 mM Na_5Cit3.06<	1.08 mM H ₂ Cit+0.04 mM Na ₃ Cit3.073.093.103.113.093.123.131.92 mM H ₂ Cit+0.31 mM Na ₃ Cit5.075.095.125.125.055.095.130.55 mM Na ₃ HPO ₄ +1.25 1mM NaH ₂ PO6.046.086.156.186.036.013.131.36 mM Na ₃ HPO ₄ +1.25 1mM NaH ₂ PO7.097.127.187.247.087.017.152.11 mM Na ₃ HPO ₄ +1.02 mM NaH ₂ PO6.096.016.166.046.013.167.01Aqueous buffer20%2.2%2.4%6.046.044.024.024.021.08 m M ₂ Cit+0.41 mM Na ₃ Cit3.173.183.213.233.213.233.211.09 m M ₂ Cit+0.42 mM Na ₃ Cit3.173.183.213.233.213.233.213.231.09 m M ₃ Cit+0.41 mM Na ₃ Cit3.173.183.213.233.213.233.213.233.231.09 m M ₃ Cit+0.42 mM Na ₃ Cit3.173.183.213.233.233.233.233.233.231.09 m M ₃ Cit+0.42 mM Na ₄ Cit3.133.143.123.143.143.143.143.143.141.09 m M ₃ Cit+0.42 mM Na ₄ Cit3.133.143.143.143.143.143.143.141.01 m Na ₄ HPO ₄ +1.25 mM Na ₄ Dit3.143.143.143.143.143.143.141.10 m Na ₄ HPO ₄ +1.25 mM Na ₄ Dit3.143.143.143.14<	1.08 mM H,Cit+0.04 mM Na,Cit3.073.093.003.103.153.093.123.103.141.92 mM H,Cit+0.31 mM Na,Cit4.024.034.074.114.024.064.034.070.74 mM H,Cit+0.26 mM Na,Cit5.095.095.125.125.125.106.186.186.186.186.186.186.186.186.196.196.197.127.137.147.1	1.08 mM H,Cit+0.04 mM NayCit3.073.093.093.103.103.113.143.121.92 mM H,Cit+0.31 mM NayCit4.024.034.074.114.024.064.084.014.070.74 mM H,Cit+0.26 mM NayEit5.075.075.086.0	1.08 mM H,Girle 04 mM Na,Gir3.073.093.073.093.073.093.013.093.01	108 mM Highen Mangenia500			

TABLE 1 ^S_WpH^W and ^S_SpH^S in the mobile phase of high-performance liquid chromatography.

TABLE 2 Parameter results of pentapeptides at 25°C.

Pentapeptide sequence		8.0%			10.0%			12.0%			14.0%	
	k_{HA}	k_A	pК _a	k_{HA}	k_A	pK _a	k_{HA}	k_A	pК _a	k_{HA}	k_A	pК _a
APNPT	52.5	1.8	1.0	40.9	1.2	0.9	13.5	0.9	1.3	4.5	0.7	1.9
DPNPT	10.8	0.9	2.6	6.8	0.6	2.6	4.7	0.5	2.6	3.2	0.4	2.7
EPNPT	15.7	0.9	2.4	9.2	0.5	2.5	6.2	0.3	2.5	4.0	0.2	2.6
GPNPT	19.9	1.4	1.4	18.0	1.0	1.2	6.2	0.8	1.7	82.8	0.6	0.1
HPNPT	69.6	3.4	0.9	41.4	2.1	0.9	32.0	1.4	0.8	20.8	1.0	0.8
KPNPT	37.0	1.0	0.9	24.4	0.7	0.9	90.0	0.6	0.1	12.1	0.4	0.9
NPNPT	30.0	1.5	1.2	27.7	1.0	1.0	11.3	0.7	1.3	88.8	0.5	0.1
PPNPT	90.4	3.2	1.0	53.8	2.1	1.0	17.6	1.4	1.4	27.1	1.0	0.9
RPNPT	74.5	1.9	1.0	53.6	1.3	0.9	33.4	1.0	0.9	22.0	0.7	0.9
SPNPT	13.4	1.5	1.8	19.4	1.0	1.2	8.5	0.7	1.5	91.3	0.5	0.1
TPNPT	62.7	2.5	1.1	47.2	1.6	1.0	37.4	1.1	0.9	10.6	0.8	1.4
NPNPA	20.1	1.5	1.4	17.9	1.0	1.2	6.1	0.7	1.7	81.4	0.6	0.1
NPNPD	5.5	0.5	2.5	3.5	0.3	2.6	2.4	0.2	2.7	1.7	0.1	2.9
NPNPE	9.5	0.6	2.5	6.0	0.4	2.5	4.0	0.3	2.5	2.7	0.2	2.7
NPNPG	29.5	1.1	0.8	109.0	0.8	0.0	48.4	0.6	0.2	1.7	0.4	2.4
NPNPH	31.8	1.2	0.8	92.8	0.8	0.1	61.4	0.6	0.1	92.0	0.4	-0.2
NPNPK	28.0	0.8	0.8	85.4	0.5	0.1	18.9	0.4	0.6	65.0	0.3	-0.1
NPNPN	98.5	0.9	0.0	44.0	0.6	0.2	10.9	0.4	0.7	4.6	0.3	1.0
NPNPQ	10.1	1.0	1.4	83.7	0.7	0.1	2.2	0.5	2.2	1.4	0.3	2.7
NPNPR	44.4	1.4	0.9	27.0	1.0	0.9	95.3	0.7	0.1	53.7	0.5	0.2
NPNPS	10.4	0.9	1.4	91.3	0.6	0.1	2.3	0.4	2.3	1.5	0.3	2.6
Pentapeptide sequence		20.0%			22.0%			24.0%			26.0%	
	k_{HA}	k_A	pК _a									
HGRFA	47.7	3.2	0.9	42.0	2.4	0.8	1.5	4.4	6.8	1.2	3.7	6.9
HGRFD	57.7	1.2	0.9	42.3	0.9	0.9	29.1	0.7	0.9	21.9	0.6	0.9
HGRFE	55.0	1.1	0.9	39.9	0.9	0.9	27.3	0.7	0.9	25.0	0.6	0.8
HGRFG	41.0	2.5	0.9	29.8	2.0	0.9	1.2	3.6	6.8	1.0	2.9	6.8
HGRFH	1.4	9.1	6.7	1.0	6.7	6.7	0.8	5.4	6.8	0.6	4.3	6.8
HGRFK	1.2	5.3	6.6	1.0	4.1	6.6	0.7	3.7	6.6	0.5	3.3	6.6
HGRFN	1.4	4.5	6.7	1.1	3.4	6.7	0.8	2.8	6.7	0.7	2.3	6.8
HGRFQ	35.0	2.1	0.8	1.3	3.7	6.8	1.0	3.0	6.8	0.8	2.4	6.8
HGRFR	1.6	7.8	6.6	1.2	5.9	6.6	0.8	5.2	6.6	0.6	4.6	6.6
HGRFS	34.0	2.1	0.9	25.2	1.6	0.9	1.0	3.0	6.7	0.8	2.5	6.7
HGRFT	53.0	2.8	0.9	38.2	2.2	0.9	29.0	1.8	0.8	1.1	3.3	6.8
AGRFG	40.2	2.1	0.9	30.5	1.7	0.9	26.8	1.4	0.8	24.8	1.2	0.7
DGRFG	42.5	1.5	0.9	32.4	1.2	0.9	28.9	1.0	0.8	23.0	0.8	0.8
EGRFG	49.3	1.5	0.9	37.0	1.2	0.9	26.9	1.0	0.9	25.7	0.8	0.8

Pentapeptide sequence		20.0%			22.0%			24.0%			26.0%	
	k_{HA}	k_A	pК _a									
GGRFG	32.2	1.7	0.9	29.8	1.4	0.8	26.0	1.2	0.7	85.3	1.1	0.0
KGRFG	38.0	2.3	0.9	1.4	4.8	6.5	1.0	4.5	6.6	0.7	4.2	6.6
NGRFG	38.0	1.6	0.9	35.0	1.3	0.8	25.2	1.1	0.8	110.0	0.9	0.0
PGRFG	2.8	8.7	7.0	2.2	6.9	7.0	30.2	2.0	0.8	1.3	4.9	7.0
QGRFG	44.2	1.8	0.9	40.7	1.5	0.8	30.1	1.2	0.8	23.6	1.0	0.8
RGRFG	45.9	1.6	1.0	40.5	1.3	0.9	32.0	1.2	0.8	23.5	1.0	0.8
SGRFG	38.6	1.6	0.9	35.7	1.3	0.8	25.8	1.1	0.8	20.4	0.9	0.8
TGRFG	41.0	2.1	0.9	30.7	1.7	0.9	26.6	1.4	0.8	24.5	1.2	0.7
VGRFG	4.6	15.0	7.1	3.6	11.2	7.0	2.6	8.7	7.0	2.0	6.9	7.0
NPNPC	21.9	0.4	1.5	17.9	0.4	1.3	13.1	0.3	1.2	3.4	0.2	1.9
NPNPI	55.2	2.6	1.1	40.8	2.0	1.1	45.0	1.6	0.9	10.2	1.2	1.7
NPNPM	11.8	1.2	1.6	15.6	1.0	1.2	6.2	0.7	1.7	90.8	0.6	0.1
NPNPP	2.9	0.4	2.5	2.1	0.3	2.6	1.5	0.2	2.8	1.2	0.2	2.8
NPNPV	17.3	1.0	1.2	8.4	0.8	1.5	99.1	0.6	0.1	2.7	0.5	2.3
NPNPY	18.7	1.5	1.4	19.2	1.1	1.2	10.5	0.9	1.4	4.0	0.7	2.1
CPNPT	2.4	0.5	2.3	1.7	0.4	2.5	1.3	0.3	2.6	1.0	0.2	2.9
IPNPT	52.3	2.5	0.9	35.4	1.9	0.9	30.9	1.4	0.8	22.1	1.1	0.8
LPNPT	59.0	3.0	1.0	48.1	2.3	0.9	34.4	1.7	0.9	30.2	1.3	0.8
MPNPT	39.9	1.7	0.9	34.0	1.3	0.8	24.7	1.0	0.8	80.9	0.8	0.1
QPNPT	2.0	0.5	4.5	1.4	0.3	5.2	1.1	0.2	5.4	0.9	0.1	5.6
VPNPT	27.0	1.0	0.8	83.5	0.8	0.1	50.1	0.6	0.2	2.2	0.5	1.9
YPNPT	50.9	2.3	0.9	33.8	1.7	0.9	29.5	1.2	0.8	20.9	0.9	0.8

TABLE 2 (Continued) Parameter results of pentapeptides at 25°C.

optimized to achieve satisfactory separation of mixtures and symmetric peak shapes. Furthermore, an increasing number of studies have reported the combined effect of two or more factors on the retention time (Phyo et al., 2018; Biancolillo et al., 2020; Kaczmarski and Chutkowski, 2021; Yilmaz, 2021). Comprehensive models that consider the influence of different chromatographic conditions are more accurate in predicting the retention times of compounds. However, previous studies have generally predicted the chromatographic retention or lipophilicity by using the quantitative structure-retention relationship (QSRR) models (Yang X. et al., 2020; Fouad et al., 2022; Xu et al., 2023). The QSRR models mainly focus on the molecular descriptors of the solutes, with less emphasis on the influence of different chromatographic conditions. Recently, models based on empirical or semiempirical equations and thermodynamic properties have rarely been reported to investigate the simultaneous effect of diverse chromatographic conditions on retention.

Herein, this study aims to provide multiparameter models that combine the effects of pH, temperature (T), organic

modifier composition (φ), and polarity (P_m^N) to predict the retention factors of amphoteric compounds (pentapeptides) under different chromatographic conditions in RP-HPLC. Initially, the retention factors of 57 pentapeptides under seven mobile phase pH values, three column temperatures, and four methanol compositions were measured *via* RP-HPLC. Then, the multiparameter models derived from the sigmoidal function, Van't Hoff equation, and empirical formula between log k and the solvent polarity or solvent composition were built. Finally, the multiparameter models were evaluated by comparing the agreement between experimental k-values and predicted k-values.

2 Materials and methods

2.1 Chemicals

RP-HPLC-grade methanol was purchased from Fisher Scientific, and trifluoroacetic acid (TFA) was purchased from

TABLE 3 Parameter results of pentapeptides at 35°C.

Pentapeptide sequence		8.0%			10.0%			12.0%			14.0%	
	k_{HA}	k_A	pК _a									
APNPT	49.0	1.7	0.9	33.0	1.2	0.9	15.5	0.8	1.1	87.5	0.6	0.1
DPNPT	8.0	0.8	2.6	5.3	0.6	2.6	3.5	0.3	2.8	2.5	0.2	2.8
EPNPT	10.9	0.8	2.5	7.1	0.5	2.5	4.6	0.3	2.6	3.2	0.2	2.6
GPNPT	17.7	1.3	1.3	25.2	0.9	0.9	93.6	0.7	0.1	2.8	0.5	2.0
HPNPT	49.0	2.7	0.9	31.0	1.7	0.9	24.4	1.1	0.8	72.4	0.8	0.1
KPNPT	30.5	0.9	0.9	25.0	0.7	0.8	21.2	0.5	0.7	87.5	0.4	-0.1
NPNPT	16.1	1.3	1.4	26.1	0.9	0.9	99.3	0.6	0.1	3.4	0.4	1.8
PPNPT	66.7	2.9	1.0	50.8	1.9	0.9	32.3	1.3	0.9	25.6	1.0	0.8
RPNPT	53.8	1.6	1.0	40.7	1.1	0.9	25.6	0.8	0.9	99.2	0.5	0.1
SPNPT	10.3	1.3	1.8	13.2	0.9	1.3	4.1	0.6	2.0	3.8	0.5	1.7
TPNPT	57.9	2.2	1.0	18.8	1.5	1.4	29.7	1.0	0.9	6.6	0.7	1.6
NPNPA	13.6	1.3	1.5	10.5	0.9	1.4	93.8	0.7	0.1	2.5	0.5	2.2
NPNPD	4.0	0.4	2.6	2.7	0.2	2.7	1.9	0.2	2.7	1.5	0.1	2.6
NPNPE	6.5	0.5	2.6	4.3	0.3	2.6	3.0	0.2	2.6	2.1	0.1	2.7
NPNPG	23.1	0.9	0.8	60.0	0.6	0.2	5.9	0.4	1.2	6.3	0.3	1.0
NPNPH	24.7	0.9	0.8	76.4	0.6	0.1	86.3	0.4	-0.1	79.0	0.3	-0.2
NPNPK	100.9	0.7	0.1	19.9	0.4	0.7	80.2	0.3	-0.1	61.0	0.2	-0.1
NPNPN	63.5	0.7	0.1	2.0	0.5	2.0	5.4	0.3	1.0	1.9	0.2	1.5
NPNPQ	20.9	0.9	0.8	52.7	0.6	0.2	1.9	0.4	2.1	5.3	0.3	1.0
NPNPR	32.2	1.1	0.9	25.8	0.7	0.8	78.0	0.5	0.1	83.5	0.4	-0.1
NPNPS	12.6	0.8	1.1	58.5	0.5	0.2	11.2	0.4	0.8	5.7	0.3	1.0
Pentapeptide sequence		20.0%			22.0%			24.0%			26.0%	
	k_{HA}	k_A	pК _a									
HGRFA	37.2	2.4	0.9	27.4	1.9	0.9	1.2	3.3	6.7	1.0	2.7	6.8
HGRFD	41.3	0.8	0.9	30.7	0.7	0.9	22.8	0.6	0.9	12.1	0.5	1.1
HGRFE	41.0	0.8	0.9	30.2	0.7	0.9	23.0	0.5	0.9	12.2	0.4	1.1
HGRFG	31.8	1.8	0.9	29.3	1.4	0.8	25.9	1.2	0.7	96.0	0.9	0.0
HGRFH	1.1	6.1	6.7	0.8	4.7	6.7	0.6	3.6	6.7	0.5	2.8	6.7
HGRFK	1.0	3.6	6.6	0.8	3.0	6.6	0.6	2.6	6.6	0.5	2.2	6.6
HGRFN	1.0	3.2	6.7	0.8	2.5	6.7	0.6	2.0	6.7	0.5	1.6	6.7
HGRFQ	1.2	3.4	6.7	0.9	2.7	6.7	0.7	2.1	6.8	54.9	0.7	0.1
HGRFR	1.1	5.0	6.5	0.9	4.3	6.6	0.6	3.7	6.6	0.5	3.0	6.6
HGRFS	31.1	1.5	0.8	23.3	1.2	0.8	100.6	0.9	0.0	59.2	0.8	0.1
HGRFT	38.8	2.1	0.9	28.2	1.7	0.9	25.4	1.3	0.8	23.0	1.1	0.7
AGRFG	31.5	1.6	0.9	29.8	1.3	0.8	22.6	1.1	0.8	104.4	0.9	0.0
DGRFG	32.7	1.1	0.9	25.5	0.9	0.9	24.7	0.7	0.8	91.3	0.6	0.1
EGRFG	37.9	1.1	0.9	29.1	0.9	0.9	22.7	0.7	0.9	102.7	0.6	0.1

Pentapeptide sequence		20.0%			22.0%			24.0%			26.0%	
	k_{HA}	k_A	pК _a									
GGRFG	31.2	1.3	0.8	24.0	1.1	0.8	104.0	0.9	0.0	65.6	0.8	0.1
KGRFG	29.7	1.7	0.9	1.0	3.7	6.6	0.7	3.2	6.6	0.6	2.9	6.6
NGRFG	36.0	1.2	0.8	27.7	1.0	0.8	21.7	0.8	0.8	78.5	0.7	0.1
PGRFG	2.1	7.6	7.0	1.6	6.2	7.0	1.3	5.1	7.0	1.0	4.1	6.9
QGRFG	41.0	1.4	0.8	31.7	1.2	0.8	24.6	1.0	0.8	93.0	0.8	0.1
RGRFG	41.6	1.2	0.9	30.2	1.1	0.9	27.5	0.9	0.8	91.8	0.8	0.1
SGRFG	30.4	1.2	0.9	28.6	1.0	0.8	22.5	0.8	0.8	76.7	0.8	0.1
TGRFG	32.2	1.6	0.9	30.1	1.3	0.8	22.7	1.1	0.8	103.3	0.9	0.0
VGRFG	3.5	12.1	6.9	2.6	9.3	6.9	2.0	7.2	6.9	1.6	5.7	6.9
NPNPC	41.6	0.4	0.9	32.7	0.3	0.8	6.0	0.2	1.5	3.7	0.2	1.6
NPNPI	17.2	2.3	1.8	34.9	1.8	1.1	12.4	1.4	1.6	16.1	1.1	1.3
NPNPM	14.4	1.0	1.3	7.3	0.8	1.6	99.0	0.6	0.1	3.8	0.5	1.8
NPNPP	2.3	0.3	2.6	1.8	0.3	2.5	1.5	0.2	2.3	1.3	0.1	2.2
NPNPV	12.6	0.9	1.3	5.2	0.7	1.8	86.2	0.6	0.1	2.6	0.4	2.2
NPNPY	8.3	1.1	1.8	10.3	0.9	1.4	5.4	0.7	1.7	3.1	0.5	2.1
CPNPT	6.9	0.5	1.1	1.8	0.4	2.0	1.2	0.3	2.3	1.0	0.2	2.3
IPNPT	41.7	2.4	0.9	29.4	1.8	0.9	30.0	1.4	0.7	20.0	1.1	0.8
LPNPT	56.0	2.9	0.9	39.4	2.2	0.9	28.2	1.6	0.9	27.0	1.3	0.8
MPNPT	38.4	1.5	0.8	28.3	1.1	0.8	20.1	0.9	0.8	74.9	0.7	0.1
QPNPT	1.6	0.4	4.8	1.2	0.2	5.3	1.0	0.2	4.4	0.8	0.2	4.4
VPNPT	21.9	1.0	0.8	70.2	0.8	0.1	42.4	0.6	0.2	12.0	0.5	0.7
YPNPT	34.9	1.8	0.9	30.0	1.3	0.8	20.8	1.0	0.8	77.4	0.7	0.1

TABLE 3 (Continued) Parameter results of pentapeptides at 35°C.

Fluka (Buchs, Switzerland). All other reagents were from Kermel (Tianjin, China); these included citric acid, sodium citrate, disodium hydrogen phosphate, and sodium dihydrogen phosphate.

The pentapeptides (HGRFG and NPNPT) were isolated from Carapax Trionycis and showed high anti-fibrosis activity (Supplementary Figure S1). The C- or N-termini of the pentapeptides of HGRFG and NPNPT were replaced with the remaining 19 amino acids to obtain the sequences of the derived pentapeptides. Then, the derived pentapeptides were synthesized by solid-phase synthesis (SPPS) and purified by RPLC. In this study, the sequences of the 57 analyzed pentapeptides are as follows: NPNPA, NPNPC, NPNPD, NPNPE, NPNPG, NPNPH, NPNPI, NPNPK, NPNPM, NPNPN, NPNPP, NPNPQ, NPNPR, NPNPS, NPNPT, NPNPV, NPNPY, APNPT, CPNPT, DPNPT, EPNPT, GPNPT, HPNPT, IPNPT, KPNPT, LPNPT, MPNPT, PPNPT, QPNPT, RPNPT, SPNPT, TPNPT, VPNPT, YPNPT, HGRFA, HGRFD, HGRFE, HGRFG, HGRFH, HGRFK, HGRFN, HGRFQ, HGRFR, HGRFS, HGRFT, AGRFG, DGRFG, EGRFG, GGRFG, KGRFG, NGRFG, PGRFG, QGRFG, RGRFG, SGRFG, TGRFG, and VGRFG.

2.2 Instruments

RP-HPLC was conducted *via* a Shimadzu Prominence LC-2030 Plus (Kyoto, Japan) instrument equipped with a SIL-20AC autosampler and two LC-20AD pumps. An SPD-20AV dual-wavelength detector at 215 nm and 254 nm was used to detect the pentapeptides. Instrument control, data acquisition, and processing were performed with LabSolutions software for RP-HPLC. A Shimadzu Shim-pack GIST C18 4.6 × 250 mm i. d., 5 μ m particle size column was used as the stationary phase and was stable within the pH range of 1–10.

A PHS-25 pH meter purchased from INESA (Shanghai, China) was used to measure the pH values, combined with an E-201F-type composite electrode. Potassium hydrogen phthalate, mixed phosphate, and sodium tetraborate from INESA (Shanghai, China) were used for electrode calibration.

2.3 Chromatographic procedure

Mobile phases were prepared with water (A)-methanol (B) components, degassed, and mixed online. The pentapeptides

TABLE 4 Parameter results of pentapeptides at 45°C.

Pentapeptide sequence		8.0%			10.0%			12.0%			14.0%	
	k_{HA}	k_A	pК _a	k_{HA}	k_A	pК _a	k_{HA}	k_A	pК _a	k_{HA}	k_A	pК _a
APNPT	43.0	1.5	1.0	33.6	1.1	0.9	22.8	0.8	0.9	71.7	0.6	0.2
DPNPT	6.8	0.7	2.6	4.3	0.5	2.7	2.9	0.3	2.9	2.0	0.2	3.0
EPNPT	9.1	0.7	2.5	5.6	0.5	2.6	3.6	0.3	2.7	2.6	0.2	2.7
GPNPT	38.1	1.1	0.9	10.2	0.8	1.4	101.9	0.6	0.1	2.2	0.5	2.2
HPNPT	43.4	2.1	0.9	32.8	1.4	0.8	21.7	0.9	0.8	55.5	0.6	0.2
KPNPT	28.9	1.0	0.9	23.0	0.7	0.8	60.0	0.5	0.2	17.5	0.4	0.6
NPNPT	39.7	1.1	0.9	10.7	0.7	1.4	85.3	0.5	0.2	3.9	0.4	1.6
PPNPT	66.0	2.7	1.0	41.0	1.8	1.0	38.0	1.3	0.8	24.6	1.0	0.8
RPNPT	56.0	1.5	0.9	34.9	1.0	0.9	22.5	0.7	0.9	79.0	0.6	0.1
SPNPT	10.7	1.1	1.7	11.0	0.8	1.4	107.6	0.6	0.1	2.7	0.4	2.0
TPNPT	57.8	1.9	1.0	18.5	1.3	1.4	29.9	0.9	0.9	8.1	0.7	1.4
NPNPA	26.7	1.1	1.1	10.6	0.8	1.4	103.0	0.6	0.1	2.1	0.5	2.3
NPNPD	3.6	0.4	2.5	2.2	0.2	2.8	1.6	0.1	2.8	1.2	0.1	2.8
NPNPE	5.6	0.5	2.5	3.4	0.2	2.7	2.4	0.2	2.7	1.7	0.1	2.8
NPNPG	25.0	0.8	0.8	62.2	0.6	0.2	2.1	0.4	2.0	6.4	0.3	1.0
NPNPH	24.0	0.8	0.8	60.3	0.5	0.2	99.5	0.4	-0.2	72.6	0.3	-0.2
NPNPK	107.1	0.6	0.1	24.0	0.4	0.6	78.8	0.3	-0.1	60.2	0.2	-0.1
NPNPN	58.3	0.6	0.2	1.9	0.4	2.0	5.5	0.3	1.0	1.0	0.2	2.1
NPNPQ	5.8	0.7	1.6	2.4	0.5	2.1	7.2	0.4	1.0	5.3	0.3	1.0
NPNPR	29.1	1.0	0.9	102.8	0.7	0.1	22.9	0.5	0.6	19.0	0.4	0.5
NPNPS	15.8	0.7	1.0	2.9	0.5	1.9	2.4	0.3	1.8	5.6	0.3	1.0
Pentapeptide sequence		20.0%			22.0%			24.0%			26.0%	
	k_{HA}	k_A	pК _a	k_{HA}	k_A	pК _a	k_{HA}	k_A	pК _a	k_{HA}	k_A	pК _a
HGRFA	1.3	4.6	6.7	1.1	3.6	6.7	0.8	2.9	6.7	0.7	2.4	6.8
HGRFD	28.7	0.7	0.9	22.6	0.6	0.9	11.9	0.5	1.1	11.2	0.4	1.0
HGRFE	29.6	0.7	0.9	23.1	0.6	0.9	12.1	0.5	1.1	11.3	0.4	1.0
HGRFG	1.1	3.6	6.7	0.9	2.9	6.8	26.2	1.0	0.8	0.5	1.9	6.8
HGRFH	0.7	4.9	6.7	0.6	3.8	6.7	0.5	3.0	6.7	0.4	2.5	6.8
HGRFK	0.6	5.0	6.7	0.5	4.1	6.8	0.4	3.4	6.8	0.3	3.0	6.8
HGRFN	0.7	2.9	6.7	0.5	2.3	6.7	0.4	1.9	6.7	0.3	1.6	6.8
HGRFQ	0.8	3.0	6.7	0.6	2.4	6.7	0.5	2.0	6.8	0.4	1.6	6.9
HGRFR	0.8	3.8	6.9	0.5	4.4	6.6	0.4	3.9	6.7	0.3	3.4	6.7
HGRFS	0.9	3.0	6.7	0.7	2.5	6.8	0.5	2.0	6.7	0.4	1.7	6.8
HGRFT	1.2	4.2	6.7	1.0	3.3	6.7	0.8	2.7	6.8	0.6	2.2	6.8
AGRFG	29.5	1.4	0.8	24.4	1.1	0.8	0.8	2.1	6.9	0.6	1.8	6.9
DGRFG	27.0	0.9	0.9	22.7	0.7	0.9	104.4	0.6	0.1	11.6	0.5	1.0
EGRFG	31.8	0.8	0.9	25.4	0.7	0.9	21.5	0.5	0.9	25.7	0.4	0.7
·	1				1					(Continu	ad an fall	wing nage

Pentapeptide sequence		20.0%			22.0%			24.0%			26.0%	
	k_{HA}	k_A	pK _a	k_{HA}	k_A	pK _a	k_{HA}	k_A	pК _a	k_{HA}	k_A	рК _а
GGRFG	24.0	1.1	0.8	90.0	0.9	0.1	70.4	0.8	0.1	41.4	0.7	0.2
KGRFG	0.8	5.4	6.8	0.6	4.9	6.8	0.4	4.5	6.8	0.3	4.2	6.8
NGRFG	29.0	1.0	0.8	22.7	0.9	0.8	87.5	0.7	0.1	52.8	0.6	0.2
PGRFG	1.4	7.6	7.0	1.1	6.1	7.0	0.9	5.0	7.0	0.7	4.2	7.0
QGRFG	32.6	1.2	0.8	26.6	1.0	0.8	20.6	0.9	0.8	69.6	0.8	0.1
RGRFG	35.0	1.1	0.8	27.5	0.9	0.8	96.5	0.8	0.1	68.0	0.7	0.1
SGRFG	29.9	1.0	0.8	23.5	0.9	0.8	90.9	0.7	0.1	55.3	0.6	0.2
TGRFG	31.0	1.3	0.8	24.4	1.1	0.8	90.1	0.9	0.1	64.9	0.8	0.1
VGRFG	2.6	10.6	6.9	1.9	8.0	6.8	1.4	6.3	6.8	1.1	5.1	6.9
NPNPC	15.8	0.3	1.3	12.6	0.3	1.2	2.7	0.2	2.0	1.8	0.2	2.0
NPNPI	44.8	2.0	1.1	39.0	1.5	1.1	11.1	1.2	1.7	33.9	1.0	0.9
NPNPM	13.4	0.9	1.3	7.9	0.7	1.5	90.5	0.5	0.2	3.4	0.4	1.8
NPNPP	2.0	0.3	2.6	2.6	0.3	1.8	1.3	0.2	2.4	6.0	0.2	1.0
NPNPV	10.8	0.8	1.4	5.1	0.6	1.8	2.8	0.5	2.2	8.1	0.4	1.2
NPNPY	14.0	0.9	1.3	7.0	0.7	1.6	90.7	0.5	0.2	2.0	0.4	2.5
CPNPT	2.0	0.5	2.0	6.1	0.3	1.1	2.0	0.2	1.7	9.1	0.2	0.7
IPNPT	39.7	2.1	0.9	28.2	1.7	0.9	27.0	1.2	0.8	105.2	1.0	0.0
LPNPT	44.1	2.5	1.0	39.5	1.9	0.9	14.3	1.4	1.3	26.0	1.1	0.8
MPNPT	36.5	1.3	0.8	27.9	1.0	0.8	97.8	0.8	0.1	59.3	0.6	0.2
QPNPT	1.5	0.4	4.6	1.1	0.3	5.1	0.9	0.2	5.4	0.7	0.2	5.5
VPNPT	22.2	0.9	0.8	78.6	0.7	0.1	41.5	0.5	0.3	7.1	0.4	1.0
YPNPT	38.1	1.4	0.8	28.5	1.0	0.8	96.3	0.8	0.1	45.5	0.6	0.3

TABLE 4 (Continued) Parameter results of pentapeptides at 45°C.



FIGURE 1

Fitting of the experimental retention and different pH values of VGRFG at 35°C and 20% methanol (A), HGRFA at 25°C and 24% methanol (B), and QPNPT at 45°C and 26% methanol (C).



were analyzed under isocratic elution of organic solvent B. The analysis procedures were, respectively, as follows: a: 8–14 v/v (increment 2 v/v) (NPNPA, NPNPD, NPNPE, NPNPG, NPNPH, NPNPK, NPNPN, NPNPQ, NPNPR, NPNPS, NPNPT, APNPT, DPNPT, EPNPT, GPNPT, HPNPT, KPNPT, PPNPT, RPNPT, SPNPT, and TPNPT); b: 20–26 v/v (increment 2 v/v) (HGRFA, HGRFD, HGRFE, HGRFG, HGRFH, HGRFK, HGRFN, HGRFQ, HGRFR, HGRFS, HGRFT, AGRFG, DGRFG, EGRFG, GGRFG, KGRFG, NGRFG, PGRFG, QGRFG, RGRFG, SGRFG, TGRFG, VGRFG, NPNPC, NPNPI, NPNPM, NPNPP, NPNPV, NPNPY, CPNPT, IPNPT, LPNPT, MPNPT, QPNPT, VPNPT, and YPNPT). The retention times were separately obtained at temperatures of 25°C, 35°C, and 45°C. The aqueous phase was prepared at 25°C by diluting stock solutions of buffer salt.

The parameters ${}^{W}_{W}pH^{W}$ and ${}^{S}_{W}pH^{W}$ of the RP-HPLC mobile phase were associated with the chromatographic retention of ionizable compounds through their thermodynamic acid-base constants in the methanol-water mixture. The ${}^{W}_{W}pH^{W}$ and ${}^{S}_{W}pH^{W}$ values were recorded before and after mixing water with the organic phase after the electrode was calibrated with the pH calibration solution at the working temperature, and the ${}^{S}_{S}pH^{S}$ values were calculated according to Eqs 2, 3. The ${}^{S}_{W}pH^{W}$ and ${}^{S}_{S}pH^{S}$ values at different temperatures and organic modifier compositions are shown in Table 1. All pH values are named according to IUPAC nomenclature.

The solutes were initially dissolved in pure water at a concentration lower than 1 mg/ml and then filtered through a 0.45 μ m nylon mobile phase filter. The flow rate of the chromatographic system was maintained at 1.0 ml/min, and the injection volume was 10 μ L.

2.4 Data statistics and analysis

Both non-linear regressions of the chromatographic retention factor k with pH or other parameters in the multiparameter equation and linear regression were performed using MATLAB R2019a (Version 9.6.0; MathWorks, Natick, MA, USA).

3 Theory

3.1 Influence of pH

The theoretical sigmoidal function of pH and retention factor k derived from chromatographic theory has been widely used for



FIGURE 3

Plots of the logk_{HA}, logk_A, and pK_a of 8%-14% methanol acid pentapeptides at 35°C (A), 20%-26% methanol basic pentapeptides at 25°C (B), 20%-26% methanol basic pentapeptides at 35°C (C), 8%-14% methanol neutral pentapeptides at 25°C (D), and 20%-26% methanol neutral pentapeptides at 25°C (E), 20%–26% methanol basic pentapeptides at 45°C (F) as a function of φ or P_m^N

ionizable compounds (Konçe et al., 2019). Previous studies have verified the wide applicability of ionizable compounds in chromatographic analysis (Yang et al., 2018; Soriano-Meseguer et al., 2019). Thus,

according to Equation 1, the acid-base equilibrium determined by the acidity constant Ka, i.e., the retention factor of the monoprotic acid solute, depends on the pH of the mobile phase.

parameter a b c d a b c d e f AINIT 1.51 4.50 0.80 1.238 0.42 0.40 0.398 1.099 0.400 1.912 0.900 0.900 DPNYT 0.66 1.123 0.29 5.938 2.60 0.000 0.98 1.912 0.400 2.767 6.66 EPNYT 0.66 1.237 0.20 5.275 0.60 1.35 6.13 0.400 5.275 0.66 2.115 0.30 5.275 0.66 1.121 0.135 0.145	φ	8%								1(0%		
DPNPT 0.633 11.23 0.279 53.8 2.601 0.000 1.88 1.402 4.001 4.001 2.781 5.611 GPNPT 1.62 1.327 0.279 5.381 0.408 2.113 0.879 11.30 4.001 4.001 7.022 HNNT 1.371 11.615 0.083 11.401 0.000 1.365 6.135 0.406 9.655 0.713 5.761 KPNPT 1.319 6.110 0.026 0.371 0.300 1.366 2.111 0.113 6.116 4.060 4.061 4.118 4.089 4.010 1.366 2.111 0.112 5.401 4.010 4.010 4.010 1.012 1.012 1.012 1.011 0.010 1.029 4.112 1.038 4.011 4.51 4.010 4.011 1.011 4.010 4.011 1.011 1.012 1.011 1.011 1.011 1.011 1.011 1.011 1.011 1.011 1.011 1.011	parameter		b	с	d				b	с	d		f
EINIT 0.62 1.3.20 0.2.79 5.9.8 2.648 5.870 0.498 1.132 0.301 0.000 2.875 1.400 -7.621 GRNFT 1.376 11.463 0.081 5.561 0.490 0.000 1.353 6.135 0.208 5.275 0.713 5.705 MNNFT 1.39 6.110 0.022 0.713 0.703 1.546 1.215 0.104 6.410 1.400 0.000 1.405 0.112 0.714 0.414 5.71 0.412 5.780 0.413 5.600 0.413 5.610 1.312 1.381 0.214 6.314 6.314 6.310 0.370 0.412 5.780 0.410 5.781 0.411 7.720 1.613 5.61 0.71 1.343 0.208 5.275 1.620 0.403 SPNPT 0.711 2.669 0.413 5.680 0.575 0.600 0.527 1.629 0.412 1.419 0.403 0.527 1.629 0.411	APNPT	1.541	4.630	0.093	4.208	0.942	0.810	1.398	5.076	0.009	1.912	0.900	0.000
GPNPT 1.7% 1.383 0.068 5.561 0.409 26.113 0.879 11.300 0.208 5.275 1.400 7.692 HPNPT 1.371 11.615 0.083 11.401 0.900 0.000 1.355 6.135 0.000 9.655 0.713 5.611 RPNPT 1.319 6.110 0.026 0.71 0.900 0.000 1.446 1.213 0.000 0.655 5.700 NPNPT 1.544 -3.999 0.112 7.230 0.613 5.616 1.312 0.318 0.414 5.410 0.400 0.800 1.421 1.338 -0.310 5.71 0.172 5.740 0.813 5.611 1.312 1.338 -0.318 5.71 0.137 0.129 0.413 5.61 0.313 5.611 1.312 0.308 5.31 1.420 0.401 5.71 1.421 7.901 0.457 1.432 0.464 1.331 0.328 5.275 1.632 1.619 1.337	DPNPT	0.583	11.223	-0.279	5.938	2.600	0.000	0.398	10.949	-0.370	4.007	2.787	-5.061
HPNPT 1.371 11.615 0.683 11.401 0.900 0.000 1.355 6.135 0.900 0.600 1.354 0.900 0.600 1.354 1.131 0.100 0.600 1.346 1.213 0.155 0.500 0.555 0.557 NPNT 1.548 -3.390 -0.112 7.320 0.755 1.3563 0.656 1.113 0.416 6.373 0.412 0.340 0.600 1.496 6.173 0.12 3.740 0.422 0.600 SPNT 0.557 7.18 0.412 7.320 1.613 5.661 1.312 0.208 5.275 1.62 -0.301 SPNT 0.567 0.416 0.483 0.887 14.372 0.494 0.208 5.275 1.681 -0.420 NPNPD 0.316 0.415 0.416 0.857 1.437 0.428 5.257 1.681 -0.416 NPNPD 0.316 0.416 0.488 2.581 0.431 0.428	EPNPT	0.662	13.270	-0.279	5.938	2.645	-5.870	0.496	11.822	-0.301	0.000	2.687	-5.061
KPNPT1.3196.1010.0260.3710.9000.0001.3461.2130.1550.0000.6555.791NPNPT1.548-3.990.1127.200.7551.35630.65621.1150.3168.2101.6890.101PPNPT1.6247.9830.4344.8001.0000.0001.4966.1730.1223.7400.9420.810GNNPT0.5775.8710.1127.3201.6135.6611.31210.8980.1416.3510.5011.324SNPT1.7112.6990.1436.4810.8555.6610.5211.3230.0194.7111.792.3483NPNPD0.3514.4670.1230.89714.320.6821.31310.2085.711.6301.111NPNPD0.4514.4680.6892.580.8100.1231.31310.2071.3310.3271.3021.0191NPNPD0.4534.4677.050.8000.0001.3281.4290.4037.340.4291.0191NPNPG1.7214.480.2677.750.8000.0001.5891.0390.5881.6261.332NPNPG1.2127.740.229.230.8000.0002.583.3090.581.3320.6911.3420.591NPNPG1.2137.040.4299.230.6000.2353.30190.581.5921.0230.	GPNPT	1.796	-13.863	-0.068	5.561	0.408	26.113	0.879	11.300	-0.208	5.275	1.400	-7.692
NNNT1.548-3.99-0.1127.3200.7531.5600.6552.1115-0.3168.2101.689-1.943PNNT1.6247.9880.444.6801.0000.0001.4986.1730.1223.2400.0010.001RNNT1.5997.4180.4115.8000.8135.6111.3121.038-0.1285.751.621.5231.6211.521SNNT1.7112.0490.1436.4810.8555.700.6822.8230.0104.7911.2022.424NNNA1.4554.6870.1127.3200.8971.4720.6411.1170.9041.3373.321.931NNNA1.4554.6870.1127.3200.8971.4320.4311.4170.4933.333.433.4321.931NNNA1.4551.4870.1127.3200.8971.4370.4911.1170.9491.333.431.949NNNA1.4571.4840.582.580.6011.1500.5331.6331.6340.431.949NNNR1.4174.4890.4267.9750.8000.6011.4160.5351.6391.5331.6421.529NNNR1.4574.4490.4480.5970.4451.6491.6491.6491.6491.6491.6491.649NNNR1.4570.4580.4790.4580.4690.4691.649 <t< td=""><td>HPNPT</td><td>1.371</td><td>11.615</td><td>0.083</td><td>11.401</td><td>0.900</td><td>0.000</td><td>1.355</td><td>6.135</td><td>-0.060</td><td>9.655</td><td>0.713</td><td>5.061</td></t<>	HPNPT	1.371	11.615	0.083	11.401	0.900	0.000	1.355	6.135	-0.060	9.655	0.713	5.061
PPNT1.6247.9880.4340.4081.0000.0001.4986.1730.1720.7410.4020.000RPNT1.5997.4180.4145.8000.4135.6161.51210.3930.4146.5451.0200.000SPNT0.4730.4730.4730.4730.4730.4730.4730.4730.4730.4730.4730.4730.4730.4730.4740.4730.4740.4730.4740.4710.4740.4710.474 <t< td=""><td>KPNPT</td><td>1.319</td><td>6.110</td><td>-0.026</td><td>0.371</td><td>0.900</td><td>0.000</td><td>1.346</td><td>1.213</td><td>-0.155</td><td>0.000</td><td>0.655</td><td>5.870</td></t<>	KPNPT	1.319	6.110	-0.026	0.371	0.900	0.000	1.346	1.213	-0.155	0.000	0.655	5.870
RPNT 1.559 7.418 0.044 5.800 0.813 5.061 1.312 10.388 0.141 6.354 0.900 0.0001 SPNTT 0.878 5.871 0.112 7.320 1.613 5.061 0.731 1.3.84 0.208 5.275 1.632 1.011 TPNTT 1.111 2.069 0.143 6.481 0.855 5.870 0.682 2.3.83 0.019 4.791 1.979 2.3.482 NPNTA 1.435 -4.467 -0.112 7.320 0.897 1.437 0.694 1.3.39 -0.28 5.275 1.680 -1.171 NPNTA 1.435 -4.467 -0.12 7.320 0.897 -0.810 0.123 1.147 -0.331 3.032 -10.311 NPNTA 1.274 4.498 -0.267 7.705 0.800 0.000 1.412 1.428 0.33 3.309 -0.33 4.649 -3.162 NPNTM 1.211 7.074 -0.322 0.923 </td <td>NPNPT</td> <td>1.548</td> <td>-3.969</td> <td>-0.112</td> <td>7.320</td> <td>0.755</td> <td>13.563</td> <td>0.656</td> <td>21.115</td> <td>-0.316</td> <td>8.210</td> <td>1.689</td> <td>-19.433</td>	NPNPT	1.548	-3.969	-0.112	7.320	0.755	13.563	0.656	21.115	-0.316	8.210	1.689	-19.433
SPNT 0.878 5.871 0.112 7.320 1.613 5.061 0.731 1.8324 0.208 5.275 1.632 0.1031 TNNT 1.111 2.069 0.143 6.481 0.855 5.870 0.682 2.823 0.019 4.711 1.979 2.3482 NNNA 1.435 4.867 0.112 7.200 0.897 1.4372 0.691 1.331 1.028 5.275 1.689 1.1171 NNNPD 0.316 10.435 0.583 5.680 0.800 0.111 0.951 1.033 1.624 2.922 1.0131 NPNPE 0.453 1.274 4.498 0.267 7.755 0.800 0.000 1.639 1.013 0.432 1.412 0.403 7.33 0.49 1.112 NPNPM 1.211 7.074 0.322 0.923 0.403 1.412 0.403 7.33 4.645 1.612 NPNPN 1.457 1.3071 0.48 8.210 <td< td=""><td>PPNPT</td><td>1.624</td><td>7.983</td><td>0.343</td><td>4.080</td><td>1.000</td><td>0.000</td><td>1.496</td><td>6.173</td><td>0.172</td><td>3.740</td><td>0.942</td><td>0.810</td></td<>	PPNPT	1.624	7.983	0.343	4.080	1.000	0.000	1.496	6.173	0.172	3.740	0.942	0.810
TrNPT1.7112.6690.1436.4810.8555.8700.6822.3.8230.0194.7911.9392.3.823NNPA1.4354.8670.1127.300.8971.4320.6041.3.910.2085.2751.6891.017NNPD0.31410.4550.5885.692.580.6100.10211.1170.95410.330.4021.0131NNPA0.45312.950.4164.6482.5880.0001.4201.4290.4037.3300.4840.101NPNPG1.2744.4990.2677.7550.8000.0001.5691.01590.5351.1340.2830.4360.4330.4681.117NPNPM1.2117.0740.3229.9330.3011.5691.01590.5351.3430.4681.33910.4531.6451.33910.4531.6451.33910.4531.6451.5491.5591.5491.5591.5491.5591.5491.5591.5591.5591.5591.5591.5591.5591.5591.5591.5591.5	RPNPT	1.559	7.418	0.044	5.800	0.813	5.061	1.312	10.398	-0.141	6.354	0.900	0.000
NPNA1.434.4670.1127.3200.8971.4320.6041.3300.2085.2751.6891.714NNPD0.31610.450.5385.682.58-0.8100.12313.65-1.03316.242.92-1.031NPNF0.45312.910.4164.6482.589-0.8100.32313.65-1.03316.242.92-1.031NPNF1.274.4980.2677.700.8000.0001.5914.42-0.4037.340.49-1.174NPNF1.217.074.3289.230.8000.0001.590.1590.55111.3420.287-5.661NPNF1.217.0744.3280.3046.720.0131.6147.9340.5881.3290.568NPNF1.45713.070.4380.3466.7210.9111.6167.9373.851.026NPNF1.45713.010.4386.731.0292.9220.9037.3790.4687.9373.891.026NPNF1.221.4160.186.271.0292.9220.4697.379.934.6531.026NPNF1.241.4160.186.292.6220.9057.7370.4164.6483.511.226NPNF1.241.4160.180.4292.6220.9597.3790.4164.483.511.567NPNF1.241.5161.516	SPNPT	0.878	5.871	-0.112	7.320	1.613	5.061	0.731	13.824	-0.208	5.275	1.632	-10.931
NPRPD0.31610.435-0.5385.6892.558-0.8100.0221.117-0.93410.3373.0.22-1.031NPRPE0.4531.2951-0.4164.6482.558-0.8100.23313.655-1.03316.2462.322-1.031NPRPG1.2744.948-0.2677.0750.8000.0001.43214.429-0.4037.3340.489-1.171NPRPH1.2117.074-0.3229.2330.8000.0001.58910.159-0.51511.3420.287-5.616NPRPK1.2813.039-0.4346.7220.91341.0931.6547.9344.6979.5334.605-105.68NPRPD0.7779.634-0.4388.2031.426-5.263-1.0767.894.6477.3973.895-102.64NPRPR1.221.041-0.1888.2031.0492.6271.0597.9734.6483.521-105.65NPRPR1.221.041-0.1888.2031.0492.6271.0597.9734.6483.521-105.65NPRPR1.221.041-0.1888.2031.0422.6271.0597.9734.0483.52-105.75NPRPR1.221.041-0.1888.2031.6261.6261.6281.7281.7373.895-102.65NPRPR1.221.0411.0280.292.6271.6901.6181.1351.626 <td< td=""><td>TPNPT</td><td>1.711</td><td>2.069</td><td>0.143</td><td>6.481</td><td>0.855</td><td>5.870</td><td>0.682</td><td>23.823</td><td>0.019</td><td>4.791</td><td>1.979</td><td>-23.482</td></td<>	TPNPT	1.711	2.069	0.143	6.481	0.855	5.870	0.682	23.823	0.019	4.791	1.979	-23.482
NPNPE0.45312.951-0.4164.6482.558-0.8100.23313.655-1.03316.2462.932-1.031NPNPG1.2744.498-0.2677.7050.8000.0001.43214.429-0.4037.3440.489-1.1741NPNPH1.2117.074-0.3229.9230.8000.0001.56910.159-0.55111.3420.287-5.061NPNPK2.858-33.933-0.3646.792-0.91341.0930.55533.019-0.5585.6891.382-30.162NPNPM1.45713.071-0.4388.2101.4265.263-1.0767.9799.9534.665-102.66NPNPR1.22410.416-0.1888.2430.9000.0002.505-2.924-0.3799.933-6.65341.295NPNPR1.4059.866-0.2795.9380.4792.267-0.9507.7790.4164.6483.521-10.501Q1.4741.4160.1888.2430.9000.0002.505-1.0750.4164.6483.521-1.026NPNPR1.4059.866-0.2795.9380.4792.262-0.9507.7970.4164.6483.521-1.026Q1.425-9.8660.2792.672-0.6531.6971.6992.3721.6911.4161.429-2.16PNPT1.4059.8660.2792.6721.68021.682 <td>NPNPA</td> <td>1.435</td> <td>-4.867</td> <td>-0.112</td> <td>7.320</td> <td>0.897</td> <td>14.372</td> <td>0.694</td> <td>13.391</td> <td>-0.208</td> <td>5.275</td> <td>1.689</td> <td>-11.741</td>	NPNPA	1.435	-4.867	-0.112	7.320	0.897	14.372	0.694	13.391	-0.208	5.275	1.689	-11.741
NPNPG1.2744.4980.2677.7050.8000.0001.43214.4290.4037.3340.489-1.174NPNPH1.2117.0740.3229.9230.8000.0001.56910.1590.5131.1340.2875.501NPNPK2.888-3.3930.3646.7920.91341.0930.5353.3019-0.5385.6991.822-30.162NPNPM1.45713.0710.4389.7950.432-10.331-1.68479.934-0.5779.5334.605-105.68NPNPQ0.7379.6340.3168.2101.4265.263-1.07679.869-0.6173.895-10.204NPNPA1.22410.160.1888.2430.9000.0002.50529.224-0.3799.034-0.53341.29NPNPS1.4059.8660.2795.9380.4792.6720.95077.3790.4164.6483.211-10.50Q1.4291.4290.490.1713.030.4772.1623.646-71.2810.3193.3932.31810.607DPNPT1.5091.6290.6211.6495.621.6495.6471.2810.3193.9392.31810.507Q1.5091.6490.6173.0300.4722.1623.646-71.2810.3193.9392.31810.507DPNFT1.5090.6490.6792.6261.6491.6	NPNPD	0.316	10.435	-0.538	5.689	2.558	-0.810	0.102	11.117	-0.954	10.337	3.032	-10.931
NPNPH1.2117.074-0.3229.9230.8000.0001.56910.159-0.55111.3420.287-5.061NPNPK2.888-33.9930.3646.722-0.91341.0930.53533.019-0.5385.6891.382-30.162NPNPN1.45713.071-0.4389.7950.432-10.931-1.68479.934-0.5779.5534.605-105.66NPNPQ0.7379.634-0.3168.2101.4265.263-1.07679.689-0.4677.9373.895-102.024NPNPR1.22410.416-0.1888.2430.9000.0002.505-2.9224-0.3799.033-0.65341.295NPNPS1.405-9.866-0.2795.9380.47922.672-0.5007.379-0.4164.6483.521-91.903\$	NPNPE	0.453	12.951	-0.416	4.648	2.558	-0.810	0.233	13.655	-1.033	16.246	2.932	-10.931
NPNPK2.858-3.3993-0.3646.792-0.91341.0930.53533.019-0.5385.6891.382-30.162NPNPN1.45713.071-0.4389.7950.432-10.931-1.68479.934-0.5779.5534.605-105.68NPNPQ0.7379.634-0.3168.2101.426-5.2631.07679.898-0.4677.9373.895-102.04NPNPR1.42410.416-0.1888.2430.9002.0022.92.40-0.3799.0332.61341.29NPNPS1.4509.866-0.2795.9380.4792.2672-0.9507.73790.4164.6483.211-1.930\$	NPNPG	1.274	4.498	-0.267	7.705	0.800	0.000	1.432	14.429	-0.403	7.334	0.489	-11.741
NPNPN1.4571.3.07-0.4389.7950.432-1.0.931-1.68479.934-0.5979.5534.605-105.68NPNPQ0.7379.634-0.3168.2101.426-5.263-1.07679.869-0.4677.9373.890-102.04NPNPR1.22410.416-0.1888.2430.0000.0002.505-29.224-0.3799.093-0.65341.29NPNPS1.405-9.866-0.2795.9380.47922.672-0.95077.379-0.4164.6483.521-91.903QQParameterabcddefabcddfabcdfAPNPT1.590 </td <td>NPNPH</td> <td>1.211</td> <td>7.074</td> <td>-0.322</td> <td>9.923</td> <td>0.800</td> <td>0.000</td> <td>1.569</td> <td>10.159</td> <td>-0.551</td> <td>11.342</td> <td>0.287</td> <td>-5.061</td>	NPNPH	1.211	7.074	-0.322	9.923	0.800	0.000	1.569	10.159	-0.551	11.342	0.287	-5.061
NPNPQ0.7379.634-0.3168.2101.426-5.263-1.07679.689-0.4677.9373.895-102.024NPNPR1.22410.16-0.1888.2430.9000.0002.505-2.9224-0.3709.093-0.65341.296NPNPS1.4059.866-0.2795.9380.47922.672-0.9507.7379-0.4164.6483.51191.903\$\phi\$1.4059.8660.2795.9380.47922.672-0.9507.73790.4164.6483.52191.903\$\phi\$1.4059.8660.2795.9380.47922.672-0.5007.73790.4164.6483.52191.903\$\phi\$1.4059.8660.2795.9380.47922.672-0.5007.73790.4164.6483.52191.903\$\phi\$1.4059.8660.2795.9380.47922.672-0.5007.73790.4164.6483.51291.933\$\phi\$1.5071.5011.5071.5011.5071.5011.5071.5011.5011.5071.5011.5011.5011.5011.5011.5011.5011.5011.5011.5011.5011.5011.5011.5011.501 </td <td>NPNPK</td> <td>2.858</td> <td>-33.993</td> <td>-0.364</td> <td>6.792</td> <td>-0.913</td> <td>41.093</td> <td>0.535</td> <td>33.019</td> <td>-0.538</td> <td>5.689</td> <td>1.382</td> <td>-30.162</td>	NPNPK	2.858	-33.993	-0.364	6.792	-0.913	41.093	0.535	33.019	-0.538	5.689	1.382	-30.162
NPNPR1.22410.416-0.1888.2430.9000.0002.505-2.9224-0.3799.093-0.65341.296NPNPS1.405-9.866-0.2795.3380.47922.672-0.50577.379-0.4164.6483.5219.19.33 φ	NPNPN	1.457	13.071	-0.438	9.795	0.432	-10.931	-1.684	79.934	-0.597	9.553	4.605	-105.668
NPNPS1.405-9.866-0.2795.9380.47922.672-0.95077.379-0.4164.6483.521-9.1903φEE	NPNPQ	0.737	9.634	-0.316	8.210	1.426	-5.263	-1.076	79.689	-0.467	7.937	3.895	-102.024
φL2L2K3K4K5K4K5K4K5K5parameterabcdcdefabcdefAPNPT1.590-12.004-0.1713.0030.43721.8623.646-71.281-0.3193.930-2.31810.0607DPNPT0.20611.649-0.84413.0243.276-16.8020.06211.198-1.13517.6723.318-15.992EPNPT0.27612.998-0.5230.0002.932-10.9310.16610.253-0.6990.0002.787-5.061GPNPT3.746-71.073-0.3646.792-2.21693.927-1.87791.647-0.4164.6485.124-121.660HPNPT1.1229.490-0.27310.5590.8000.0002.428-2.5956-0.47112.012-0.72636.032KPNPT1.26213.996-0.4164.6480.634-9.9191.880-15.067-0.3980.000-0.23923.279NPNPT3.240-52.069-0.4677.937-1.45065.385-1.40880.165-0.5385.6893.887-89.676PNPT1.1369.618-0.3468.6240.9000.0002.758-3.3394-0.3654.571-1.0584.5976SPNPT2.806-53.2250.3193.930-0.8266.6.802-1.69688.564-0.	NPNPR	1.224	10.416	-0.188	8.243	0.900	0.000	2.505	-29.224	-0.379	9.093	-0.653	41.296
Parameter a b c d e f a b c d e f APNPT 1.590 -12.004 -0.171 3.003 0.437 21.862 3.646 -71.281 -0.319 3.930 -2.318 100.607 DPNPT 0.206 11.649 -0.844 13.024 3.276 -16.802 0.062 11.198 -1.135 17.672 3.318 -15.992 EPNPT 0.276 12.998 -0.523 0.000 2.932 -10.931 0.196 10.253 -0.699 0.000 2.787 -5.061 GPNPT 3.746 -71.073 -0.364 6.792 -2.216 93.927 -1.877 91.647 -0.416 4.648 -121.66 HPNPT 1.122 9.490 -0.273 10.559 0.800 0.000 2.428 -25.956 -0.471 12.012 -0.269 3.632 3.632 NPNPT 1.262 13.996 -0.416 4.648 <t< td=""><td>NPNPS</td><td>1.405</td><td>-9.866</td><td>-0.279</td><td>5.938</td><td>0.479</td><td>22.672</td><td>-0.950</td><td>77.379</td><td>-0.416</td><td>4.648</td><td>3.521</td><td>-91.903</td></t<>	NPNPS	1.405	-9.866	-0.279	5.938	0.479	22.672	-0.950	77.379	-0.416	4.648	3.521	-91.903
APNPT 1.590 -12.004 -0.171 3.003 0.437 21.862 3.646 -71.281 -0.319 3.930 -2.318 100.607 DPNPT 0.206 11.649 -0.844 13.024 3.276 -16.802 0.062 11.198 -1.135 17.672 3.318 -15.992 EPNPT 0.276 12.998 -0.523 0.000 2.932 -10.931 0.196 10.253 -0.699 0.000 2.787 -5.661 GPNPT 3.746 -71.073 -0.364 6.792 -2.216 93.927 -1.877 91.647 -0.416 4.648 5.124 -121.660 HPNPT 1.122 9.490 -0.273 10.559 0.800 0.000 2.428 -25.956 -0.471 12.012 -0.726 36.032 KPNPT 1.262 13.996 -0.467 7.937 -1.450 65.385 -1.408 80.165 -0.538 5.689 3.887 -89.676 PNPT 2.023 -19.052 0.067 1.889 -0.011 34.413 1.340 2.327 0.000				12	2%					1-	4%		
DPNPT 0.206 11.649 -0.844 13.024 3.276 -16.802 0.062 11.198 -1.135 17.672 3.318 -15.992 EPNPT 0.276 12.998 -0.523 0.000 2.932 -10.931 0.196 10.253 -0.699 0.000 2.787 -5.061 GPNPT 3.746 -71.073 -0.364 6.792 -2.216 93.927 -1.877 91.647 -0.416 4.648 5.124 -121.660 HPNPT 1.122 9.490 -0.273 10.559 0.800 0.000 2.428 -25.956 -0.471 12.012 -0.726 36.032 KPNPT 1.262 13.996 -0.467 7.937 -1.450 65.385 -1.408 80.165 -0.538 5.689 3.887 -89.676 PNPT 3.240 -52.069 -0.467 7.937 -1.450 65.385 -1.408 80.165 -0.538 5.689 3.887 -89.676 PNPT 2.023 -1.90.52	parameter		b	с	d				b	C	d		f
Image: Constraint of the state of	APNPT	1.590	-12.004	-0.171	3.003	0.437	21.862	3.646	-71.281	-0.319	3.930	-2.318	100.607
Image: Constraint of the straint of the str	DPNPT	0.206	11.649	-0.844	13.024	3.276	-16.802	0.062	11.198	-1.135	17.672	3.318	-15.992
HPNPT 1.122 9.490 -0.273 10.559 0.800 0.000 2.428 -25.956 -0.471 12.012 -0.726 36.032 KPNPT 1.262 13.996 -0.416 4.648 0.634 -9.919 1.880 -15.067 -0.398 0.000 -0.239 23.279 NPNPT 3.240 -52.069 -0.467 7.937 -1.450 65.385 -1.408 80.165 -0.538 5.689 3.887 -89.676 PPNPT 2.023 -19.052 0.067 1.889 -0.011 34.413 1.340 2.327 0.000 0.000 0.655 5.870 RPNPT 1.136 9.618 -0.346 8.624 0.900 0.000 2.758 -33.394 -0.365 4.571 -1.058 46.964 SPNPT 2.806 -53.225 -0.319 3.930 -0.826 66.802 -1.696 88.564 -0.482 4.904 4.576 -109.109 TPNPT 1.333 5.730 <td< td=""><td>EPNPT</td><td>0.276</td><td>12.998</td><td>-0.523</td><td>0.000</td><td>2.932</td><td>-10.931</td><td>0.196</td><td>10.253</td><td>-0.699</td><td>0.000</td><td>2.787</td><td>-5.061</td></td<>	EPNPT	0.276	12.998	-0.523	0.000	2.932	-10.931	0.196	10.253	-0.699	0.000	2.787	-5.061
KPNPT 1.262 13.996 -0.416 4.648 0.634 -9.919 1.880 -15.067 -0.398 0.000 -0.239 23.279 NPNPT 3.240 -52.069 -0.467 7.937 -1.450 65.385 -1.408 80.165 -0.538 5.689 3.887 -89.676 PPNPT 2.023 -19.052 0.067 1.889 -0.011 34.413 1.340 2.327 0.000 0.000 0.655 5.870 RPNPT 1.136 9.618 -0.346 8.624 0.900 0.000 2.758 -33.394 -0.365 4.571 -1.058 46.964 SPNPT 2.806 -53.225 -0.319 3.930 -0.826 66.802 -1.696 88.564 -0.482 4.904 4.576 -109.109 TPNPT 1.333 5.730 -0.145 4.746 0.900 0.000 0.688 7.578 -0.239 3.404 1.516 -1.619	GPNPT	3.746	-71.073	-0.364	6.792	-2.216	93.927	-1.877	91.647	-0.416	4.648	5.124	-121.660
NPNPT 3.240 -52.069 -0.467 7.937 -1.450 65.385 -1.408 80.165 -0.538 5.689 3.887 -89.676 PPNPT 2.023 -19.052 0.067 1.889 -0.011 34.413 1.340 2.327 0.000 0.000 0.655 5.870 RPNPT 1.136 9.618 -0.346 8.624 0.900 0.000 2.758 -33.394 -0.365 4.571 -1.058 46.964 SPNPT 2.806 -53.225 -0.319 3.930 -0.826 66.802 -1.696 88.564 -0.482 4.904 4.576 -109.109 TPNPT 1.333 5.730 -0.145 4.746 0.900 0.000 0.688 7.578 -0.239 3.404 1.516 -1.619	HPNPT	1.122	9.490	-0.273	10.559	0.800	0.000	2.428	-25.956	-0.471	12.012	-0.726	36.032
PPNPT 2.023 -19.052 0.067 1.889 -0.011 34.413 1.340 2.327 0.000 0.000 0.655 5.870 RPNPT 1.136 9.618 -0.346 8.624 0.900 0.000 2.758 -33.394 -0.365 4.571 -1.058 46.964 SPNPT 2.806 -53.225 -0.319 3.930 -0.826 66.802 -1.696 88.564 -0.482 4.904 4.576 -109.109 TPNPT 1.333 5.730 -0.145 4.746 0.900 0.000 0.688 7.578 -0.239 3.404 1.516 -1.619	KPNPT	1.262	13.996	-0.416	4.648	0.634	-9.919	1.880	-15.067	-0.398	0.000	-0.239	23.279
RPNPT 1.136 9.618 -0.346 8.624 0.900 0.000 2.758 -33.394 -0.365 4.571 -1.058 46.964 SPNPT 2.806 -53.225 -0.319 3.930 -0.826 66.802 -1.696 88.564 -0.482 4.904 4.576 -109.109 TPNPT 1.333 5.730 -0.145 4.746 0.900 0.000 0.688 7.578 -0.239 3.404 1.516 -1.619	NPNPT	3.240	-52.069	-0.467	7.937	-1.450	65.385	-1.408	80.165	-0.538	5.689	3.887	-89.676
SPNPT 2.806 -53.225 -0.319 3.930 -0.826 66.802 -1.696 88.564 -0.482 4.904 4.576 -109.109 TPNPT 1.333 5.730 -0.145 4.746 0.900 0.000 0.688 7.578 -0.239 3.404 1.516 -1.619	PPNPT	2.023	-19.052	0.067	1.889	-0.011	34.413	1.340	2.327	0.000	0.000	0.655	5.870
TPNPT 1.333 5.730 -0.145 4.746 0.900 0.000 0.688 7.578 -0.239 3.404 1.516 -1.619	RPNPT	1.136	9.618	-0.346	8.624	0.900	0.000	2.758	-33.394	-0.365	4.571	-1.058	46.964
	SPNPT	2.806	-53.225	-0.319	3.930	-0.826	66.802	-1.696	88.564	-0.482	4.904	4.576	-109.109
NPNPA 3.766 -71.731 -0.280 3.388 -2.216 93.927 -1.933 92.633 -0.416 4.648 5.426 -128.340	TPNPT	1.333	5.730	-0.145	4.746	0.900	0.000	0.688	7.578	-0.239	3.404	1.516	-1.619
	NPNPA	3.766	-71.731	-0.280	3.388	-2.216	93.927	-1.933	92.633	-0.416	4.648	5.426	-128.340
NPNPD -0.008 9.733 -1.261 15.234 2.887 -5.061 -0.084 8.095 -1.000 0.000 2.539 7.490	NPNPD	-0.008	9.733	-1.261	15.234	2.887	-5.061	-0.084	8.095	-1.000	0.000	2.539	7.490
NPNPE 0.115 12.239 -0.954 10.337 2.932 -10.931 -0.007 11.052 -1.436 17.672 2.887 -5.061	NPNPE	0.115	12.239	-0.954	10.337	2.932	-10.931	-0.007	11.052	-1.436	17.672	2.887	-5.061

TABLE 5 Parameter results of pH and T under different mobile phase compositions.

TABLE 5 (Continued)	Parameter results	s of pH and T under	different mobile phase compositions.	
THELE S (Continued)	i araniecer resards	or prir unior i unioci	unterent mobile phase compositions.	

φ			12	2%					14	4%		
parameter		b	с	d				b	с	d		
NPNPG	-1.390	76.360	-0.653	10.337	4.142	-99.190	1.636	-33.743	-0.704	7.334	-1.026	82.186
NPNPH	2.265	-11.807	-0.653	10.337	-0.576	16.802	1.733	5.741	-0.704	7.334	-0.200	0.000
NPNPK	2.798	-36.462	-0.704	7.334	-1.113	41.093	1.735	1.909	-0.954	10.337	-0.100	0.000
NPNPN	0.306	17.503	-0.704	7.334	1.434	-17.611	-0.798	36.650	-0.954	10.337	3.345	-59.717
NPNPQ	1.268	-25.543	-0.538	5.689	-0.100	61.538	1.561	-33.940	-0.523	0.000	-1.461	99.798
NPNPR	0.772	32.043	-0.513	8.578	1.034	-25.304	0.998	21.283	-0.538	5.689	0.587	-12.753
NPNPS	0.794	-6.502	-0.631	6.323	0.497	37.449	1.581	-33.647	-0.523	0.000	-1.316	93.927
			2	0%					2	2%		
parameter		b	с	d				b	C	d		
HGRFA	-1.307	80.053	0.727	-6.964	11.737	-293.522	-1.440	81.555	0.650	-8.090	11.882	-299.393
HGRFD	1.110	16.525	-0.460	13.272	0.900	0.000	1.037	14.904	-0.438	9.795	0.900	0.000
HGRFE	1.164	14.650	-0.405	11.054	0.900	0.000	1.087	12.992	-0.438	9.795	0.900	0.000
HGRFG	-1.387	80.417	0.611	-6.859	11.737	-293.522	-1.370	76.981	0.513	-6.912	11.866	-297.773
HGRFH	-0.477	16.082	0.356	15.012	6.700	0.000	-0.471	12.012	0.277	13.711	6.700	0.000
HGRFK	-0.529	15.875	0.580	2.641	6.787	-5.061	-0.619	16.019	0.534	1.098	6.974	-10.121
HGRFN	-0.501	16.418	0.211	10.855	6.700	0.000	-0.678	18.449	0.137	9.672	6.700	0.000
HGRFQ	-2.370	94.907	0.733	-9.534	15.239	-346.356	-0.606	18.287	0.138	10.622	6.555	5.870
HGRFR	-0.453	16.552	0.197	17.369	7.103	-14.372	-0.704	20.253	0.453	7.560	6.600	0.000
HGRFS	-1.438	80.133	0.527	-6.656	11.679	-292.713	-1.526	79.036	0.494	-8.797	11.866	-297.773
HGRFT	-1.428	84.350	0.704	-7.900	11.737	-293.522	-1.450	81.131	0.607	-8.005	11.737	-293.522
AGRFG	1.292	7.659	-0.075	9.868	0.713	5.061	1.297	4.986	-0.190	10.511	0.655	5.870
DGRFG	1.194	10.893	-0.316	12.318	0.900	0.000	1.162	8.662	-0.431	12.858	0.900	0.000
EGRFG	1.271	10.561	-0.412	14.907	0.900	0.000	1.203	9.112	-0.431	12.858	0.900	0.000
GGRFG	1.261	6.571	-0.190	10.511	0.655	5.870	2.317	-23.532	-0.273	10.559	-0.508	35.425
KGRFG	-1.615	85.720	0.978	-17.695	11.924	-298.583	-0.626	19.806	0.633	0.462	7.118	-15.992
NGRFG	1.347	6.131	-0.250	11.342	0.655	5.870	1.134	10.339	-0.250	9.005	0.800	0.000
PGRFG	-0.188	16.246	0.796	3.446	7.000	0.000	-0.300	16.354	0.712	3.085	7.000	0.000
QGRFG	1.380	6.955	-0.137	9.795	0.655	5.870	1.202	10.227	-0.209	9.696	0.800	0.000
RGRFG	1.417	6.305	-0.172	9.247	0.568	10.931	1.220	9.540	-0.226	8.669	0.713	5.061
SGRFG	1.319	6.453	-0.250	11.342	0.713	5.061	1.158	9.970	-0.250	9.005	0.800	0.000
TGRFG	1.325	6.994	-0.135	11.497	0.713	5.061	1.296	5.117	-0.190	10.511	0.655	5.870
VGRFG	0.131	13.501	0.840	8.386	6.611	11.741	-0.044	15.190	0.729	8.049	6.568	10.931
NPNPC	1.237	4.919	-0.631	6.323	0.779	14.980	1.119	5.598	-0.704	7.334	0.824	9.109
NPNPI	1.279	8.689	0.171	6.198	1.505	-5.668	1.535	1.541	0.041	6.693	1.100	0.000
NPNPM	1.225	-3.495	-0.200	6.964	0.866	17.611	0.450	17.625	-0.346	8.624	1.992	-18.421
NPNPP	0.103	8.982	-0.704	7.334	2.745	-5.870	0.457	-4.152	-0.523	0.000	1.047	41.296
NPNPV	0.776	11.470	-0.208	5.275	1.632	-10.931	0.399	12.654	-0.364	6.792	2.234	-17.611

TABLE 5 (Continued) Parameter results of pH and T under different mobile phase compositions.

φ			2	.0%					2	2%		
parameter		b	с	d				b	с	d		
NPNPY	0.833	9.218	-0.316	12.318	1.445	1.822	0.308	24.366	-0.376	10.640	2.063	-21.862
CPNPT	0.498	0.293	-0.301	0.000	1.045	24.899	1.282	-28.282	-0.631	6.323	-0.405	74.899
IPNPT	1.438	6.855	0.246	3.976	0.900	0.000	1.318	5.651	0.175	2.635	0.900	0.000
LPNPT	1.522	6.581	0.321	4.126	0.942	0.810	1.472	5.031	0.196	4.355	0.900	0.000
MPNPT	1.519	2.092	-0.019	6.336	0.655	5.870	1.325	4.991	-0.141	6.354	0.800	0.000
QPNPT	0.011	7.108	-0.538	5.689	4.861	-7.490	-0.088	5.842	-0.625	1.426	5.071	4.251
VPNPT	1.220	5.038	-0.085	2.316	0.800	0.000	1.829	1.939	-0.205	2.935	0.100	0.000
YPNPT	1.377	7.693	-0.103	11.773	0.713	5.061	1.361	4.168	-0.268	12.606	0.655	5.870
			2	4%					2	6%		
parameter		b	с	d				b	с	d		
HGRFA	-0.390	14.601	0.233	10.174	6.555	5.870	-0.404	12.487	0.138	10.622	6.655	5.870
HGRFD	0.677	20.511	-0.467	7.937	1.274	-10.121	0.647	16.824	-0.597	9.553	1.203	-6.68
HGRFE	0.733	18.486	-0.513	8.578	1.274	-10.121	0.573	19.975	-0.653	10.337	1.347	-12.55
HGRFG	3.354	-78.572	-0.759	32.016	-7.942	353.036	0.585	-0.816	-0.175	13.408	2.863	55.06
HGRFH	-0.551	11.342	0.153	14.344	6.555	5.870	-0.597	9.553	0.086	13.428	6.742	0.810
HGRFK	-0.648	12.842	0.411	3.099	6.974	-10.121	-0.716	11.227	0.339	3.521	6.974	-10.12
HGRFN	-0.732	16.246	0.048	9.706	6.700	0.000	-0.927	19.806	-0.024	9.252	6.742	0.810
HGRFQ	-0.652	16.489	0.058	10.166	6.800	0.000	0.404	0.364	-0.259	13.244	3.108	49.19
HGRFR	-0.732	16.246	0.397	7.520	6.787	-5.061	-0.830	15.875	0.310	8.147	6.787	-5.06
HGRFS	0.597	-0.981	-0.155	13.145	2.821	54.251	0.423	0.099	-0.201	12.483	3.066	48.38
HGRFT	-1.484	79.378	0.502	-7.767	12.011	-303.644	0.314	2.631	-0.087	12.775	3.268	49.39
AGRFG	-1.464	77.778	0.415	-8.063	12.197	-308.704	-1.264	76.742	0.336	-7.900	11.879	-308.0
DGRFG	2.464	-27.677	-0.504	12.481	-0.508	35.425	1.153	10.196	-0.551	11.342	0.768	-4.45
EGRFG	1.205	5.522	-0.652	16.489	0.900	0.000	1.758	-4.872	-0.732	16.246	0.208	10.72
GGRFG	2.572	-26.768	-0.322	9.923	-0.826	36.032	1.278	16.812	-0.405	11.054	0.432	-10.93
KGRFG	-0.833	21.393	0.567	1.199	6.974	-10.121	-0.881	19.164	0.530	1.302	6.974	-10.12
NGRFG	2.374	-26.833	-0.405	11.054	-0.508	35.425	1.361	17.318	-0.438	9.795	0.432	-10.93
PGRFG	-2.162	88.276	1.280	-23.430	15.974	-363.968	-0.454	14.528	0.520	4.015	6.942	0.810
QGRFG	1.120	9.045	-0.200	6.964	0.800	0.000	2.595	-28.593	-0.237	5.689	-0.913	41.09
RGRFG	2.363	-23.727	-0.322	9.923	-0.508	35.425	2.576	-28.144	-0.346	8.624	-0.913	41.09
SGRFG	2.399	-27.198	-0.405	11.054	-0.508	35.425	2.452	-26.575	-0.404	9.326	-0.726	36.03
TGRFG	2.375	-26.256	-0.273	10.559	-0.508	35.425	2.541	-26.471	-0.322	9.923	-0.826	36.03
VGRFG	-0.153	14.528	0.630	7.760	6.568	10.931	-0.240	13.924	0.546	7.316	6.755	5.870
NPNPC	-0.361	37.458	-0.954	10.337	2.868	-42.915	0.037	13.681	-0.699	0.000	1.913	-2.63
NPNPI	0.193	35.296	-0.063	6.792	2.800	-46.154	2.098	-28.002	-0.091	4.313	-0.026	43.72
NPNPM	3.664	-68.663	-0.467	7.937	-2.029	88.866	-1.505	83.357	-0.597	9.553	4.261	-99.79
NPNPP	0.060	3.145	-0.699	0.000	1.763	24.291	1.405	-35.654	-0.873	2.437	-0.911	95.95

φ			2	4%					2	6%		
parameter		b	с	d				b	С	d		
NPNPV	-0.933	78.877	-0.370	4.007	4.024	-106.275	1.313	-24.013	-0.538	5.689	0.187	56.478
NPNPY	2.604	-45.051	-0.586	13.802	-0.668	58.300	-0.024	16.131	-0.694	13.483	2.847	-20.243
CPNPT	0.443	-9.186	-0.852	8.912	0.745	47.976	1.792	-48.534	-0.699	0.000	-1.558	116.194
IPNPT	1.373	3.069	0.021	3.388	0.742	0.810	2.585	-33.942	-0.036	2.095	-0.695	40.486
LPNPT	0.774	19.991	0.058	4.480	1.647	-20.243	1.330	3.685	-0.022	3.672	0.800	0.000
MPNPT	2.458	-29.520	-0.208	5.275	-0.508	35.425	1.637	7.098	-0.364	6.792	0.287	-5.061
QPNPT	-0.145	4.746	-0.699	0.000	4.821	8.097	-0.279	5.938	-0.263	-17.672	4.718	14.777
VPNPT	1.505	4.726	-0.370	4.007	0.387	-5.061	1.720	-31.716	-0.482	4.904	-0.476	55.263
YPNPT	2.342	-24.773	-0.296	9.553	-0.508	35.425	2.281	-21.702	-0.438	9.795	-0.539	30.972

TABLE 5 (Continued) Parameter results of pH and T under different mobile phase compositions.



neutral pentapeptides (**D**) in the six-parameter model of T and pH.

$$k = \frac{k_{HA} + k_A 10^{(pH-pKa)}}{1 + 10^{(pH-pKa)}},$$
(1)

where k_{HA} and k_A represent the limiting retention factors of the protonated and dissociated forms of the analyte, respectively. The

pKa is the acid-base equilibrium constant of the solute at a given mobile phase composition and temperature. Most important, the pH here is ${}_{S}^{S}pH^{S}$ because this has been widely studied and it has been found that the fitting ability of this equation can be ensured only when the pH and pKa correspond to the real values. ${}_{S}^{S}pH^{S}$ can be

TABLE 6 Parameter results of fitting with $\boldsymbol{\phi}$ and pH at different temperatures.

TABLE 6 Parameter results of fitting with φ and pH at differen T 25'C										5°C			45°C							
Parameter	A	В	С	D	E	F	A	В	С	D	E	F	A	В	С	D	E	F		
APNPT	3.30	-18.41	0.78	-6.78	-0.43	15.50	1.35	2.14	0.84	-7.66	1.96	-11.00	1.32	2.49	0.71	-6.66	2.07	-12.00		
DPNPT	1.72	-8.73	0.38	-5.68	2.46	1.50	1.57	-8.48	0.77	-10.54	2.26	4.00	1.53	-8.83	0.60	-9.27	2.03	7.00		
EPNPT	1.96	-9.76	0.81	-10.91	2.17	3.00	1.75	-8.93	0.71	-10.14	2.33	2.00	1.67	-9.12	0.60	-9.27	2.24	3.50		
GPNPT	0.55	6.97	0.62	-6.00	2.97	-17.00	2.27	-9.16	0.65	-6.77	0.36	6.50	2.73	-13.58	0.49	-5.76	-0.28	13.00		
HPNPT	2.50	-8.43	1.22	-8.85	1.07	-2.00	1.38	2.02	1.13	-8.87	2.05	-12.50	1.48	0.70	1.05	-9.12	1.83	-10.50		
KPNPT	1.99	-4.45	0.50	-6.30	1.14	-4.00	0.82	6.51	0.44	-6.01	2.28	-15.50	1.59	-1.19	0.52	-6.70	1.45	-7.50		
NPNPT	0.92	5.12	0.80	-7.93	2.55	-15.00	2.08	-7.23	0.80	-8.56	0.83	2.00	2.45	-10.61	0.60	-7.32	0.53	4.50		
PPNPT	2.72	-10.27	1.17	-8.46	1.02	0.50	2.41	-7.22	1.07	-7.76	1.23	-3.00	2.33	-6.59	0.99	-7.18	1.34	-4.00		
RPNPT	2.60	-8.97	0.84	-7.07	1.09	-1.50	1.36	2.98	0.87	-8.27	2.21	-13.50	1.49	1.29	0.69	-6.74	2.02	-12.00		
SPNPT	0.15	10.71	0.80	-7.93	3.79	-24.00	1.83	-9.03	0.67	-7.11	1.48	2.00	1.58	-4.02	0.62	-7.21	1.52	-2.00		
TPNPT	2.85	-12.08	1.04	-8.24	0.66	4.00	2.78	-13.15	1.01	-8.34	0.51	6.50	2.65	-11.76	0.85	-7.30	0.79	3.50		
NPNPA	0.57	6.77	0.69	-6.74	2.97	-17.00	1.82	-6.28	0.65	-6.77	0.86	4.00	2.48	-11.63	0.49	-5.76	-0.04	11.50		
NPNPD	1.41	-8.47	0.62	-11.37	1.96	6.50	1.16	-7.15	0.29	-9.03	2.65	0.00	1.16	-7.85	0.38	-10.54	2.23	4.50		
NPNPE	1.70	-9.08	0.40	-7.78	2.22	3.00	1.46	-8.14	0.62	-11.37	2.46	1.50	1.41	-8.52	0.48	-10.48	2.18	4.50		
NPNPG	3.59	-20.35	0.62	-7.21	-1.90	25.00	2.66	-13.50	0.59	-8.04	-0.08	8.00	2.87	-16.23	0.49	-7.27	-0.32	12.00		
NPNPH	1.14	6.02	0.70	-7.78	1.85	-15.00	0.92	7.84	0.59	-8.04	1.91	-16.00	0.84	8.30	0.43	-6.87	2.02	-17.00		
NPNPK	1.37	2.21	0.43	-6.87	1.56	-11.00	1.78	-0.25	0.52	-8.79	0.92	-7.00	1.90	-1.17	0.40	-7.78	0.84	-6.50		
NPNPN	3.86	-22.99	0.59	-8.04	-1.45	17.50	3.06	-20.70	0.60	-9.27	-0.61	16.00	3.36	-24.18	0.40	-7.78	-1.26	23.50		
NPNPQ	3.14	-20.77	0.70	-8.57	-1.70	30.00	2.79	-16.15	0.59	-8.04	-0.35	12.50	0.48	1.80	0.32	-6.00	3.02	-14.50		
NPNPR	1.26	3.98	0.75	-7.48	2.12	-14.50	0.74	8.61	0.60	-7.32	2.46	-18.50	2.19	-6.04	0.52	-6.70	0.91	-3.50		
NPNPS	3.15	-20.61	0.59	-8.04	-1.59	29.00	2.13	-8.76	0.43	-6.87	0.61	1.50	1.49	-7.17	0.35	-6.63	1.48	-0.50		
HGRFA	8.07	-31.23	0.00	2.26	-23.75	120.00	7.75	-30.35	-0.05	1.97	-23.20	117.50	1.06	-4.72	1.60	-4.71	6.38	1.50		
HGRFD	3.19	-7.12	1.08	-5.06	0.90	0.00	3.37	-8.64	0.59	-3.40	0.26	3.00	2.96	-7.52	0.66	-4.04	0.40	2.50		
HGRFE	2.91	-5.96	0.94	-4.49	1.22	-1.50	3.34	-8.49	0.97	-5.25	0.26	3.00	3.01	-7.68	0.66	-4.04	0.40	2.50		
HGRFG	7.96	-31.17	-0.09	2.24	-23.29	118.00	0.00	6.93	1.22	-4.85	3.82	-14.00	-0.22	2.18	1.81	-6.48	11.83	-28.50		
HGRFH	1.34	-6.00	2.02	-5.35	6.29	2.00	1.18	-5.76	1.92	-5.65	6.70	0.00	0.66	-4.04	1.66	-4.90	6.38	1.50		
HGRFK	1.40	-6.48	1.37	-3.31	6.60	0.00	1.03	-5.14	1.26	-3.52	6.60	0.00	0.79	-5.00	1.44	-3.73	6.43	1.50		
HGRFN	1.18	-5.21	1.60	-4.79	6.38	1.50	1.03	-5.14	1.50	-5.00	6.70	0.00	1.04	-6.00	1.31	-4.29	6.38	1.50		
HGRFQ	6.18	-25.18	0.34	0.41	-15.40	90.00	-5.20	24.36	2.78	-10.84	27.73	-98.50	0.88	-4.91	1.37	-4.49	5.97	3.50		
HGRFR	1.66	-7.27	1.61	-3.71	6.60	0.00	1.25	-6.02	1.43	-3.65	6.23	1.50	1.25	-6.87	0.81	-0.99	7.30	-2.50		
HGRFS	7.94	-31.43	-0.22	2.50	-22.88	116.00	-0.04	7.37	1.11	-4.72	3.76	-14.50	1.16	-6.01	1.31	-4.18	6.52	1.00		
HGRFT	7.15	-25.84	0.24	0.63	-17.89	88.00	2.29	-3.63	1.28	-4.79	1.63	-3.50	1.09	-5.00	1.55	-4.65	6.29	2.00		
AGRFG	2.27	-3.43	1.13	-4.07	1.63	-3.50	-0.07	7.21	1.02	-4.11	3.73	-13.50	8.18	-32.80	-0.51	3.04	-24.21	122.00		
DGRFG	2.47	-4.25	1.07	-4.49	1.31	-2.00	0.05	6.62	0.94	-4.49	3.55	-12.50	1.97	-2.19	0.78	-4.16	1.30	-2.50		
EGRFG	2.66	-4.94	1.07	-4.49	1.22	-1.50	0.23	5.95	0.94	-4.49	3.46	-12.00	1.81	-1.75	0.97	-5.25	1.54	-3.00		
GGRFG	0.19	6.05	0.85	-3.17	3.82	-14.00	-0.17	8.03	0.83	-3.60	3.76	-14.50	1.01	3.02	0.67	-3.20	2.37	-9.00		
KGRFG	6.55	-26.75	-0.29	3.78	-14.63	86.00	6.30	-26.19	-0.29	3.16	-14.49	85.50	1.36	-7.27	1.09	-1.82	6.80	0.00		
NGRFG	0.21	6.21	1.02	-4.11	3.73	-13.50	0.51	4.55	0.88	-4.00	3.04	-10.50	0.05	6.83	0.79	-3.87	3.35	-12.50		
PGRFG	0.44	0.69	2.17	-6.43	12.58	-31.00	1.38	-5.28	1.77	-4.44	7.32	-1.50	1.14	-4.95	1.74	-4.30	7.00	0.00		
QGRFG	2.62	-4.74	1.12	-4.31	1.17	-1.50	0.52	4.78	0.96	-4.04	3.04	-10.50	0.51	4.39	0.64	-2.87	3.04	-10.50		
RGRFG	2.66	-4.87	0.84	-3.24	1.68	-3.50	0.49	4.95	0.70	-3.08	3.55	-12.50	0.08	7.05	0.67	-3.20	3.67	-14.00		

т				5°C						5°C			45°C							
Parameter				D						D						D		F		
SGRFG	2.58	-4.86	1.02	-4.11	1.17	-1.50	0.28	5.51	0.69	-3.13	3.41	-12.00	0.04	6.94	0.79	-3.87	3.35	-12.50		
TGRFG	2.32	-3.67	1.13	-4.07	1.63	-3.50	-0.02	6.98	1.02	-4.11	3.73	-13.50	-0.10	7.65	0.83	-3.60	3.67	-14.00		
VGRFG	1.89	-6.13	2.29	-5.61	7.37	-1.50	1.67	-5.67	2.17	-5.46	6.90	0.00	1.66	-6.27	2.07	-5.28	6.85	0.00		
NPNPC	4.01	-12.81	0.68	-5.14	0.21	5.50	5.59	-19.45	0.66	-5.40	-2.02	14.00	4.77	-17.50	0.20	-3.52	-1.71	14.50		
NPNPI	3.98	-10.79	1.52	-5.52	-0.64	8.00	1.89	-2.68	1.43	-5.35	2.60	-5.00	2.50	-4.54	1.29	-5.00	1.20	0.00		
NPNPM	-1.34	11.29	1.14	-5.29	5.75	-20.00	1.84	-3.02	1.03	-5.14	1.20	0.00	1.97	-3.64	1.16	-6.01	0.97	1.00		
NPNPP	1.75	-6.48	0.66	-5.40	1.41	5.50	1.17	-4.11	1.16	-8.04	4.01	-7.00	-0.90	5.65	0.20	-3.52	6.78	-21.00		
NPNPV	2.70	-6.74	1.03	-5.14	-0.91	9.50	2.00	-4.18	1.09	-5.62	0.20	5.00	1.50	-3.18	0.88	-4.91	1.88	-1.00		
NPNPY	3.66	-11.36	1.25	-5.40	-1.12	11.50	2.59	-7.82	1.19	-5.68	0.37	6.00	2.70	-7.11	1.16	-6.01	-1.13	11.00		
CPNPT	1.63	-6.29	1.04	-6.59	0.39	9.50	3.39	-13.46	1.04	-6.59	-2.56	19.50	-1.13	7.45	1.02	-6.85	5.17	-16.50		
IPNPT	2.88	-5.91	1.60	-6.01	1.31	-2.00	2.56	-4.74	1.50	-5.63	1.40	-2.50	0.19	6.25	1.44	-5.59	3.87	-14.00		
LPNPT	2.79	-5.09	1.70	-6.10	1.59	-3.00	2.82	-5.48	1.64	-5.92	1.22	-1.50	2.75	-5.65	1.60	-6.01	1.23	-1.00		
MPNPT	0.71	3.91	1.32	-5.48	3.41	-12.00	0.72	3.61	1.25	-5.40	3.04	-10.50	0.34	5.89	1.22	-5.52	3.35	-12.50		
QPNPT	1.43	-5.73	1.98	-11.37	1.15	17.50	1.18	-4.91	0.41	-4.52	7.14	-10.50	1.25	-5.40	0.66	-5.40	1.70	15.00		
VPNPT	5.36	-17.44	1.03	-5.14	-3.16	17.00	2.63	-5.01	1.03	-5.14	0.68	-1.00	3.45	-8.81	1.16	-6.01	-0.37	4.00		
YPNPT	2.91	-6.09	1.74	-6.87	1.31	-2.00	0.55	4.39	1.60	-6.72	3.41	-12.00	0.80	3.80	1.34	-6.00	3.03	-11.00		

TABLE 6 (Continued) Parameter results of fitting with ϕ and pH at different temperatures.

calculated using Equation 2, as follows (Alvarez-Segura et al., 2019; Soriano-Meseguer et al., 2019):

$${}^{s}_{S}pH^{s} = {}^{s}_{W}pH^{W} - \delta, \qquad (2)$$

where the empirical formula could be used to estimate δ from solvent composition as follows:

$$\delta = \frac{0.09\varphi_{MeOH} - 0.11\varphi_{MeOH}^2}{1 - 3.15\varphi_{MeOH} + 3.51\varphi_{MeOH}^2 - 1.35\varphi_{MeOH}^3},\tag{3}$$

where φ_{MeOH} is the volume fraction of methanol in the mixed mobile phase.

3.2 Influence of temperature

For a reversible process of chromatographic analysis, the dissociation of the analyte and buffer and the solute migration during retention, which could be affected by the column temperature change, are applicable to the Van't Hoff equation (Faisal et al., 2018; Marchetti et al., 2019; Yuan et al., 2020) as follows:

$$logk = -\frac{\Delta H^0}{2.3RT} + \frac{\Delta S^0}{2.3R} + log \Phi, \qquad (4)$$

where ΔH^0 and ΔS^0 represent the D-values of enthalpy and entropy, respectively, when the solute is transferred from the mobile to the stationary phase; R is the gas constant; and Φ represents the phase ratio. Here, we assume that the enthalpy and entropy of this equilibrium process are definite constants in the studied temperature range and that the phase ratio Φ is free of the effect.

Similarly, for the reversible process $HA \rightleftharpoons H^+ + A^-$, the acid dissociation constant pKa is the negative logarithm of the equilibrium constant Ka; thus, the correlation between pKa and temperature is described by the Van't Hoff equation as follows:

$$pK_a = \frac{\Delta H_a^0}{2.3RT} - \frac{\Delta S_a^0}{2.3R} - \log\Phi,$$
(5)

where ΔH_a^0 and ΔS_a^0 are the changes in enthalpy and entropy caused by solute dissociation, respectively.

3.3 Simultaneous influence of pH and temperature

Introducing Eqs 4, and 5 into Eq. 1 produces the following equation:

$$k = \frac{10^{a+\frac{b}{T}} + 10^{c+\frac{d}{T}} 10^{\left(pH-e-\frac{f}{T}\right)}}{1 + 10^{\left(pH-e-\frac{f}{T}\right)}},\tag{6}$$

where the fitting parameter includes the thermodynamic quantities related to the dissociation and transformation of the analyzed compound, i.e., the function composed of these quantities: $a = (\frac{\Delta S_{HA}^0}{2.3R} + \log \Phi)$, $b = \frac{-\Delta H_{HA}^0}{2.3R}$, $c = (\frac{\Delta S_{A}^0}{2.3R} + \log \Phi)$, $d = \frac{-\Delta H_{A}^0}{2.3R}$, $e = (\frac{\Delta S_{A}^0}{2.3R} + \log \Phi)$, $f = \frac{-\Delta H_{A}^0}{2.3R}$, and the subscripts HA and A apart represent the protonation and deprotonation forms of acid–base solutes.

т	25°C								35	°C			45°C						
Parameter	А		С	D	E		А	В	С	D	E		A		С	D	Е	F	
APNPT	-12.02	15.42	-5.05	5.90	12.16	-12.62	2.62	-1.20	-5.70	6.60	-6.83	8.79	3.07	-1.71	-4.97	5.73	-7.79	9.91	
DPNPT	-5.75	7.54	-4.54	4.98	3.67	-1.21	-5.67	7.31	-8.11	8.95	5.60	-3.36	-6.02	7.63	-7.25	7.92	7.96	-5.98	
EPNPT	-6.40	8.45	-8.54	9.45	4.76	-2.62	-5.87	7.70	-7.94	8.73	4.00	-1.68	-6.14	7.90	-7.25	7.92	5.29	-3.08	
GPNPT	5.87	-5.28	-4.53	5.20	-10.83	13.83	-4.58	6.78	-5.14	5.85	4.54	-4.02	-8.64	11.45	-4.46	5.00	10.34	-10.65	
HPNPT	-4.74	7.32	-6.37	7.67	-0.60	1.68	2.64	-1.20	-6.47	7.68	-8.11	10.19	1.76	-0.23	-6.71	7.84	-6.82	8.69	
KPNPT	-1.60	3.59	-4.88	5.44	-2.52	3.74	5.91	-5.06	-4.68	5.17	-10.47	12.80	0.70	0.89	-5.22	5.81	-5.10	6.64	
NPNPT	4.79	-3.84	-5.98	6.85	-9.74	12.34	-3.11	5.10	-6.47	7.34	1.21	-0.19	-6.56	9.09	-5.71	6.39	4.33	-3.83	
PPNPT	-6.21	9.05	-6.05	7.30	1.64	-0.65	-3.73	6.19	-5.59	6.73	-1.36	2.62	-3.36	5.75	-5.17	6.23	-2.00	3.36	
RPNPT	-4.99	7.66	-5.21	6.11	-0.28	1.40	3.46	-2.06	-6.15	7.09	-8.87	11.12	2.10	-0.54	-5.13	5.89	-7.68	9.72	
SPNPT	8.87	-8.75	-5.98	6.85	-16.42	20.37	-5.58	7.43	-5.43	6.16	2.67	-1.12	-1.12	2.62	-5.50	6.18	-1.28	2.99	
TPNPT	-7.18	10.08	-6.02	7.13	3.68	-2.99	-8.54	11.44	-6.10	7.18	6.14	-5.70	-7.55	10.32	-5.41	6.32	4.00	-3.27	
NPNPA	5.71	-5.10	-5.13	5.89	-10.83	13.83	-2.93	4.70	-5.14	5.85	3.23	-2.24	-7.04	9.55	-4.46	5.00	9.04	-9.07	
NPNPD	-5.85	7.33	-9.02	9.73	7.43	-5.51	-4.98	6.21	-7.49	7.88	2.81	-0.19	-5.61	6.84	-8.78	9.28	6.35	-4.21	
NPNPE	-6.06	7.83	-6.25	6.71	4.65	-2.43	-5.50	7.03	-9.02	9.73	3.67	-1.21	-5.91	7.40	-8.64	9.24	6.14	-4.02	
NPNPG	-12.14	15.64	-5.50	6.18	17.13	-18.88	-8.13	10.79	-6.29	6.95	5.96	-5.98	-10.41	13.33	-5.68	6.23	9.38	-9.72	
NPNPH	6.67	-5.64	-5.95	6.71	-11.41	13.46	8.07	-7.30	-6.29	6.95	-12.26	14.39	8.31	-7.60	-5.51	6.01	-12.91	15.14	
NPNPK	3.59	-2.29	-5.51	6.01	-8.19	9.91	0.76	1.14	-7.04	7.65	-4.20	5.05	0.18	1.84	-6.25	6.71	-3.99	4.77	
NPNPN	-15.62	19.65	-6.29	6.95	13.29	-14.86	-15.87	19.31	-7.25	7.92	14.85	-15.89	-18.34	22.07	-6.25	6.71	20.11	-21.78	
NPNPQ	-13.03	16.09	-6.56	7.33	21.59	-23.18	-10.40	13.23	-6.29	6.95	9.81	-10.19	1.58	-1.04	-4.83	5.20	-8.49	11.50	
NPNPR	4.42	-3.16	-5.62	6.43	-10.03	12.24	7.80	-7.09	-5.71	6.39	-13.04	15.61	-2.16	4.27	-5.22	5.81	-2.94	4.02	
NPNPS	-12.86	15.92	-6.29	6.95	20.95	-22.43	-4.26	6.29	-5.51	6.01	0.69	0.09	-5.54	7.23	-5.36	5.78	2.47	-1.21	
HGRFA	-18.11	26.30	1.80	-1.76	76.64	-100.75	-17.78	25.68	1.52	-1.55	75.20	-98.79	-2.95	4.05	-2.43	4.07	7.60	-1.21	
HGRFD	-2.88	6.12	-3.27	4.40	0.90	0.00	-3.90	7.32	-2.29	2.90	2.71	-2.43	-3.43	6.46	-2.76	3.45	2.53	-2.15	
HGRFE	-2.21	5.19	-2.89	3.87	0.00	1.21	-3.81	7.20	-3.45	4.45	2.71	-2.43	-3.52	6.59	-2.76	3.45	2.53	-2.15	
HGRFG	-18.26	26.37	1.73	-1.79	75.56	-99.25	5.57	-5.51	-2.91	4.17	-7.77	11.59	2.10	-2.52	-3.80	5.71	-13.97	26.64	
HGRFH	-3.80	5.20	-2.57	4.64	7.97	-1.68	-3.76	5.00	-2.90	4.86	6.70	0.00	-2.76	3.45	-2.52	4.24	7.60	-1.21	
HGRFK	-4.06	5.49	-1.50	2.92	6.60	0.00	-3.35	4.42	-1.75	3.04	6.60	0.00	-3.44	4.26	-1.76	3.24	7.79	-1.40	
HGRFN	-3.27	4.50	-2.51	4.16	7.60	-1.21	-3.35	4.42	-2.76	4.31	6.70	0.00	-4.10	5.20	-2.36	3.71	7.60	-1.21	
HGRFQ	-16.33	23.14	0.97	-0.74	66.07	-84.11	14.44	-19.43	-6.25	9.04	-52.52	79.72	-3.33	4.26	-2.45	3.87	8.87	-2.90	
HGRFR	-4.51	6.23	-1.60	3.27	6.60	0.00	-3.85	5.15	-1.66	3.12	7.59	-1.40	-4.67	6.01	0.07	0.71	4.90	2.52	
HGRFS	-18.49	26.58	1.78	-1.97	74.29	-97.57	6.09	-6.13	-2.92	4.08	-8.42	12.24	-3.96	5.17	-2.25	3.60	7.43	-0.93	
HGRFT	-14.17	21.27	0.58	-0.26	53.80	-71.21	-0.88	3.23	-2.81	4.13	-1.27	2.90	-3.14	4.26	-2.42	4.01	7.97	-1.68	
AGRFG	-0.72	3.04	-2.35	3.52	-1.27	2.90	5.72	-5.72	-2.48	3.54	-7.41	11.12	-19.34	27.64	1.96	-2.44	77.99	-102.62	
DGRFG	-1.18	3.69	-2.76	3.87	-0.37	1.68	5.29	-5.15	-2.89	3.87	-6.69	10.19	0.38	1.50	-2.80	3.62	-1.10	2.52	
EGRFG	-1.59	4.32	-2.76	3.87	0.00	1.21	4.88	-4.54	-2.89	3.87	-6.32	9.72	0.22	1.64	-3.45	4.45	-0.91	2.43	
GGRFG	5.05	-4.80	-1.88	2.77	-7.77	11.59	6.51	-6.68	-2.24	3.10	-8.42	12.24	4.19	-3.45	-2.08	2.79	-5.91	8.60	
KGRFG	-17.26	24.44	3.21	-3.64	63.09	-80.19	-17.11	24.07	2.72	-3.15	62.91	-79.91	-4.81	6.23	-0.47	1.58	6.80	0.00	

TABLE 7 Parameter results of fitting with ${\it P}_m^N$ and pH at different temperatures.

т	25°C								35	°C			45°C							
Parameter	А		С	D	E		А	В	С	D	E		А		С	D	E	F		
NGRFG	5.16	-4.87	-2.48	3.54	-7.41	11.12	4.02	-3.41	-2.53	3.44	-5.52	8.50	5.74	-5.70	-2.48	3.29	-7.16	10.56		
PGRFG	1.37	-1.07	-3.45	5.73	-15.48	28.97	-3.13	4.56	-2.01	3.82	6.10	1.21	-3.08	4.26	-1.93	3.71	7.00	0.00		
QGRFG	-1.36	3.99	-2.55	3.70	-0.19	1.40	4.22	-3.60	-2.46	3.45	-5.52	8.50	3.92	-3.32	-1.82	2.50	-5.52	8.50		
RGRFG	-1.43	4.11	-1.93	2.81	-1.35	3.08	4.32	-3.73	-1.90	2.62	-6.69	10.19	5.93	-5.86	-2.08	2.79	-8.06	11.78		
SGRFG	-1.49	4.09	-2.48	3.54	-0.19	1.40	4.68	-4.35	-2.01	2.74	-6.51	9.91	5.82	-5.79	-2.48	3.29	-7.16	10.56		
TGRFG	-0.86	3.24	-2.35	3.52	-1.27	2.90	5.58	-5.52	-2.48	3.54	-7.41	11.12	6.25	-6.36	-2.24	3.10	-8.06	11.78		
VGRFG	-3.31	5.25	-2.50	4.85	6.01	1.40	-3.18	4.90	-2.48	4.70	6.90	0.00	-3.69	5.41	-2.45	4.57	6.71	0.19		
NPNPC	-6.62	10.63	-3.57	4.24	4.38	-4.02	-10.73	16.40	-3.99	4.71	9.57	-11.59	-9.93	14.77	-2.75	2.96	10.34	-12.06		
NPNPI	-4.97	8.96	-3.18	4.75	5.79	-6.36	-0.08	1.87	-3.12	4.60	-2.20	5.05	-1.53	4.13	-3.00	4.34	1.47	-0.37		
NPNPM	7.84	-9.12	-3.35	4.53	-10.87	16.64	-0.51	2.29	-3.35	4.42	0.93	0.37	-0.83	2.72	-3.96	5.17	1.47	-0.37		
NPNPP	-3.79	5.60	-3.99	4.71	6.12	-4.77	-2.36	3.58	-5.47	6.62	-1.92	5.98	3.73	-4.60	-2.75	2.96	-11.15	18.13		
NPNPV	-2.74	5.38	-3.35	4.42	6.61	-7.38	-1.38	3.35	-3.67	4.80	4.19	-3.93	-1.64	3.34	-3.33	4.26	1.79	-0.19		
NPNPY	-5.70	9.34	-3.38	4.68	8.07	-9.07	-3.79	6.33	-3.61	4.83	4.86	-4.30	-3.01	5.63	-3.96	5.17	7.61	-8.60		
CPNPT	-3.75	5.44	-4.53	5.60	8.45	-8.13	-8.45	12.11	-4.53	5.60	14.69	-17.66	5.38	-6.64	-4.94	6.07	-9.23	14.67		
IPNPT	-2.19	5.14	-3.52	5.17	-0.37	1.68	-1.51	4.12	-3.31	4.86	-0.73	2.15	5.08	-4.78	-3.31	4.79	-7.59	11.40		
LPNPT	-1.55	4.38	-3.49	5.24	-0.99	2.62	-1.92	4.81	-3.41	5.10	0.00	1.21	-2.12	4.94	-3.52	5.17	0.46	0.75		
MPNPT	3.74	-2.95	-3.35	4.73	-6.51	9.91	3.42	-2.58	-3.38	4.68	-5.52	8.50	5.22	-4.89	-3.48	4.75	-7.16	10.56		
QPNPT	-3.50	4.99	-7.66	9.73	16.45	-15.61	-3.03	4.26	-3.67	4.22	-1.15	8.13	-3.38	4.68	-3.99	4.71	14.74	-13.27		
VPNPT	-8.30	13.35	-3.35	4.42	9.80	-12.52	-0.88	3.25	-3.35	4.42	-1.04	2.06	-3.11	6.29	-3.96	5.17	2.04	-2.06		
YPNPT	-2.33	5.31	-4.11	5.90	-0.37	1.68	3.97	-3.34	-4.12	5.78	-6.51	9.91	3.96	-3.17	-3.80	5.20	-6.25	9.35		

TABLE 7 (Continued) Parameter results of fitting with P_m^N and pH at different temperatures.

3.4 Influence of the organic modifier composition

The composition of the mobile phase is the main variable used to optimize retention and selectivity in RP-HPLC. The Soczewiński–Wachtmeister equation is commonly used to describe the relationship between k and the change in mobile phase (Flieger et al., 2020; Lin et al., 2022).

$$logk = -S\varphi + logk_W,\tag{7}$$

where $logk_W$ is the intercept and represents the retention coefficient of solute in pure water, *S* is the slope of the equation and represents the sensitivity of solute molecules to solvent strength, and φ is the volume fraction of organic modifier in the mobile phase.

Considering the influence of the polarity of the solute, stationary phase, and mobile phase on k, another linear model was proposed to accurately describe k, which represents the linear relationship between the retention rate and the polarity of the eluent (Gisbert-Alonso et al., 2021; Zhu et al., 2022); the relationship is as follows:

$$logk = (logk)_0 - p(P_m^N - P_s^N),$$
(8)

where p is the parameter describing the polarity of the solute, P_m^N and P_s^N are the standard polarity parameters of the mobile and stationary phases,

and $(logk)_0$ is the retention factor when the polarity of the mobile phase is the same as that of the stationary phase. Numerous experimental studies have shown that for a specific column and water-methanol mobile phase, the parameters of Eq. 8 can be obtained by measuring the retention rate logk of a group of solutes, where $(logk)_0$ and P_s^N are the system constants. In addition, the logk value in the model is linear with respect to P_m^N over the entire range of water-methanol mobile phase compositions (0–100%) and intersects at a common extrapolation point in the majority of cases.

For the water–methanol system, the relationship between P_m^N and φ is as follows (Zhu et al., 2022):

$$P_m^N = 1.00 - \frac{1.33\varphi}{1 + 0.47\varphi}.$$
(9)

To eliminate the limit of all *logk* and P_m^N lines that must cross at the same point, a deformation of Eq. 8 is proposed to represent the solute in the model by two descriptors (q and p) (den Uijl et al., 2021) and shown as follows:

$$logk = q + pP_m^N, \tag{10}$$

where fitting parameters concerning the solute are twice those before, which improves the accuracy of model prediction.

The mobile phase composition affects not only the retention rate but also the ionization degree of the acid-base solute, and the addition of an



Linear fitting results of the predicted k-value and experimental k-value for all pentapeptides (A), acid pentapeptides (B), basic pentapeptides (C), and neutral pentapeptides (D) in the six-parameter model of φ and pH.

and

organic solvent to the aqueous solution containing ionizable compounds changes the value of pKa. For a specific solute, the solute parameters are constant, and pKa only depends on the solvent properties or temperature, whereas when using mixed solvents (such as the mobile phase), the solvent properties and pKa change monotonically with the mobile phase composition. Therefore, the relationship between the pKa value and the solvent volume fraction can usually be expressed as follows:

$$pKa = E + F\varphi. \tag{11}$$

Similarly, the relationship between pKa and mobile phase polarity parameters can be expressed as follows:

$$pKa = E + FP_m^N. (12)$$

3.5 Simultaneous influence of pH and organic modifier composition

Based on the aforementioned analysis, combined with the model of pH and different mobile phase compositions, the sixparameter model is obtained as follows:

$$k = \frac{10^{A+BX} + 10^{C+DX} 10^{(PH-E-FX)}}{1 + 10^{(PH-E-FX)}},$$
(13)

$$logk_{HA} = A + BX, \tag{14}$$

$$logk_A = C + DX,\tag{15}$$

$$pKa = E + FX, (16)$$

where X is the variable describing the change in the mobile phase, representing φ or P_m^N in the two-parameter solvent model. Parameters A, B, C, D, E, and F of the model have a simple chemical interpretation. If X is the volume fraction φ of the organic modifier in the mobile phase, then A, C, and E are the extrapolation of the logarithm of k in acidic and alkaline forms and the *pKa* of the compound in pure water, respectively, while B, D, and F are the changes in these parameters from pure water to pure organic solvent. Similarly, if the polarity parameter P_m^N is fitted as the X variable, then A, C, and E are the reserves in a non-polar medium, and the *pKa*, and B, D, and F are the changes from the medium to pure water (defined as $P_m^N = 1$).



Linear fitting results of the predicted k-value and experimental k-value for all pentapeptides (A), acid pentapeptides (B), basic pentapeptides (C), and neutral pentapeptides (D) in the six-parameter model of P_m^N and pH.

4 Results and discussion

Small-molecular oligopeptides commonly participate in multiple physiological and pathological processes, including the transmission of signals and the regulation of immune and inflammatory responses (Yang J. et al., 2020; Gao et al., 2022). RP-HPLC is a common approach to separate small-molecular peptides by adjusting the chromatographic conditions (Liu et al., 2022). The retention factors of 57 ionizable solute derivatives of pentapeptides of NPNPT and HGRFG were determined at seven mobile pH values, four mobile phase compositions, and three column temperatures (84 data points for each solute). Some comparative chromatograms are shown in Supplementary Figure S2. We selected the pentapeptides with high polarity, similar retention, and a similar chemical structure for the model's establishment and evaluation, and the 57 pentapeptides could be divided into five groups according to the acidity or basicity of the isoelectric points and the polarity of the pentapeptides. The groups included the following: (1) 8%-14% methanol acid pentapeptides: NPNPD, NPNPE, DPNPT, and EPNPT; (2) 8%-14% methanol basic pentapeptides: NPNPK, NPNPR, KPNPT, and RPNPT; (3) 8%-14% methanol neutral pentapeptides: NPNPA, NPNPG, NPNPH, NPNPN, NPNPQ, NPNPS, APNPT, GPNPT, HPNPT, NPNPT, PPNPT, SPNPT, and TPNPT; (4) 20%–26% methanol basic pentapeptides: HGRFA, HGRFG, HGRFH, HGRFK, HGRFN, HGRFQ, HGRFR, HGRFS, HGRFT, AGRFG, GGRFG, KGRFG, NGRFG, PGRFG, QGRFG, RGRFG, SGRFG, TGRFG, and VGRFG; and (5) 20%–26% methanol neutral pentapeptides: NPNPC, NPNPI, NPNPM, NPNPP, NPNPV, NPNPY, CPNPT, IPNPT, LPNPT, MPNPT, QPNPT, VPNPT, YPNPT, HGRFD, HGRFE, DGRFG, and EGRFG.

4.1 Function of the retention factor k and pH

The pH of the mobile phase is one of the critical factors affecting the retention of compounds in chromatography due to its interference with the ionization efficiency and change in the protonation of analytes (Fan et al., 2022; Guo et al., 2023). Because the increased or decreased degree of chromatographic retention for compounds was different with the change in pH, adjusting the pH of the mobile phase was capable of separating the compounds with similar structures or confirming the absence of unrelated impurities (Fan et al., 2022; Tengattini et al., 2022).

MATLAB R2019a was used to fit the S-curve of different pH values and the experimental retention factors (k) under the same temperature and the same mobile phase composition. From Eq. 1, we obtained the parameters k_{HA} , k_A , and pK_a , as shown in Table 2 Table 3 Table 4. With the analysis of k and their corresponding pH values, the pK_a determination by the inflection point in the curve was the most established method for its calculation for a compound (Numviyimana et al., 2019). However, the retention times in RP-HPLC were generally short owing to the high polarity of solutes, leading to unsatisfactory fitting outcomes of the S-curve and predictions for inflection point pK_a . Only the pK_a -values of the acid and basic compounds from fitting were close to the software calculation results (Figure 1), consistent with the study by Rook et al. (2021) on particular sidechains.

In addition, the *k*-values calculated by the parameters k_{HA} , k_A , pK_a , and different pH values as described in Eq. 1 were fitted with 1/T, φ , and P_m^N to examine the consistency of the parameter results in the S-curve according to Eqs 4, 5, and Eq. 7 (Supplementary Tables S1, S2). For the independent variable 1/T, we obtained 1,596 correlation coefficients (\mathbb{R}^2), and 73.1% of the \mathbb{R}^2 values were greater than 0.9. For the independent variable of φ or P_m^N , 1197 \mathbb{R}^2 values were obtained, and 95.4% of the \mathbb{R}^2 values were greater than 0.9 for the independent variable. The results suggested that the parameters conformed to the relationship described in Eqs 4, 5, and Eq. 7 and further indicated the applicability of Eq. 1.

4.2 Linear relationships between k_{HA} , k_A , and pK_a with respect to 1/T

The temperature of the column could alter the density of the mobile phase, solute diffusion coefficients, and solute-stationary phase interactions and then affect the chromatographic retention (Nagase et al., 2021). The retention of solutes generally decreased as the column temperature was increased due to the accelerated molecular movement in RP-HPLC (Caltabiano et al., 2018; Idroes et al., 2020). Moreover, the k_{HA} , k_A , and pK_a of solutes were theoretically correlated with the Van't Hoff equation since they similarly represented an equilibrium state of ionization. Therefore, the relationships between $logk_{HA}$, $logk_A$, and pK_a with respect to 1/T of each compound under different mobile phase compositions were characterized, 684 R² values (Supplementary Table S3), and 60-line charts of each parameter with respect to 1/T were obtained, and some representative graphs are shown in Figure 2. In this study, the plots of $logk_{HA}$, $logk_A$, and pK_a with respect to 1/T showed linear relationships consistent with Eqs 4 and 5 for most pentapeptides, indicating that the effect of the temperature on ΔH^0 , ΔS^0 , and $\log \Phi$ could be commonly disregarded. The linear correlations of acid pentapeptides for the plots of $logk_{HA}$ vs. 1/T were the most evident, with each R² greater than 0.92 under different mobile phase compositions. The neutral pentapeptides displayed the most apparent linear correlation between $logk_A$ and 1/T, with 63% of the $R^2 > 0.9$ and 97.5% of the $R^2 > 0.75$. The linear correlations of all pentapeptides between $logk_{HA}$, $logk_A$ with respect to 1/T (43.1% of the R² > 0.9 and 56.1% of the R² > 0.9) were better than pK_a vs. 1/T (18.4% of the $\mathbb{R}^2 > 0.9$), especially for the basic pentapeptides. Furthermore, the correlations from the same group of compounds in different methanol concentrations were quite different due to the influence of the mobile phase compositions on the retention of the pentapeptides, and the linear correlations of the acid pentapeptides were slightly better than those of the neutral and basic pentapeptides.

The linear relationship between k_{HA} , k_A , and pK_a with respect to 1/T depended on the ΔH^0 , ΔS^0 , and Φ under different chromatographic conditions. However, the contributions of enthalpy and entropy to the retention process varied with the change in temperature. Enthalpy-driven effects substituted entropy-driven effects at a higher column temperature and affected the linearity of the Van't Hoff equation (Tanase et al., 2019). Moreover, the Φ value was affected by mobile phase composition and continuously increased with an increase in temperature in the range of 5–50°C (Flieger et al., 2019). A previous study reported the linear dependence of acid compounds and the non-linear behavior of basic compounds with the Van't Hoff equation (Galaon and David, 2011; Yılmaz Ortak and Cubuk Demiralay, 2019). Except for the compounds with unique structures, other compounds potentially participated in multiple interactions with the stationary phase, likely causing the deviations from the Van't Hoff equation. In summary, the $logk_{HA}$ and $logk_A$ of acidic and neutral pentapeptides followed the description of Van't Hoff equation and had apparent linear relationships with 1/T. Therefore, in this study, only a limited prediction capacity of pK_a was observed with a minimal deviation contribution to the overall retention due to the assumption of linearity.

4.3 Linear relationships between k_{HA} , k_A , and pK_a with respect to X

An appropriate composition of the mobile phase is beneficial chromatographic separation and improving for the chromatographic peak profile and efficiencies (Guo et al., 2018; Attwa et al., 2023). Adjusting the proportion of organic modifiers in the mobile phase is the most frequently used approach to achieve the separation of a series of compounds (Hong et al., 2020; Oney-Montalvo et al., 2022). Furthermore, the methanol volume fraction φ and polarity parameter P_m^N are commonly used to characterize the composition of the mobile phase. Hence, in this study, the two factors were considered simultaneously. We examined the influence of the composition of the mobile phase on $logk_{HA}$, $logk_A$, and pK_a at a constant temperature. 90-line charts of five groups and 1026 R² values (Supplementary Table S4) were obtained, and some representative charts were selected as shown in Figure 3. Identical R² values were acquired depending on the descriptor (φ or P_m^N) of the composition of the mobile phase at the same temperature; these results indicated that the relationship between φ and P_m^N could be regarded as linear in a narrow range of mobile phase compositions. Moreover, all plots of $logk_{HA}$, $logk_A$, and pK_a with respect to X exhibited linear correlations for the five groups of pentapeptides, which was consistent with Eq. 7; Eq. 10; Eqs 11, 12. More importantly, the linear model was highly suitable for describing the relationship of acid pentapeptides between $logk_{HA}$ and $logk_A$ with respect to X, with each R² greater than

0.98 or 0.9. For neutral or basic pentapeptides, the linear fitting results of $logk_A$ and X (90% of the R² > 0.9 and 79.7% of the R² > 0.9, respectively) surpassed those of $logk_{HA}$ and pK_a . Due to the influence of temperature on chromatographic retention, there were differences in the correlations of the same group of compounds at different temperatures, and the R² values were generally lower with increasing temperature. These results indicated that the two functional relationships of Eqs 11, 12 appropriate for studying the change in were the chromatographic retention of the 57 pentapeptides in the range of mobile phase compositions, and it was very difficult to compare the superiority of the independent variables of φ and P_m^N based on the current results. Specifically, $logk_{HA}$, $logk_A$, and pK_a had evident linear relationships with φ or P_m^N , and there was basically no difference regardless of whether φ or P_m^N was used as the descriptor of the composition of the mobile phase.

The coefficient $logk_W$ of the Soczewiński–Wachtmeister equation is related to not only the length of the alkyl chain in the molecule but also the latitude of the mobile phase composition (Żesławska et al., 2022). Generally, a wider range of methanol concentrations correlates with more accurate coefficients. A previous study reported the linear relationship of k and φ at 60%–80% methanol (Elmansi et al., 2019). However, in our study, the pentapeptides analyzed generally had short carbon chains with high polarity and showed low retention in RP-HPLC. No retention occurred at a high proportion of methanol, and a broad chromatographic peak with long tailing was observed at a low proportion of methanol. Hence, we used a narrow range of methanol concentrations to produce symmetrical and sharp peaks, which was crucial for accurately recording the retention time. Furthermore, we classified the pentapeptides into two groups to investigate their linear dependence at different ranges of methanol concentrations. The linear relationships need to be verified at a wider range of methanol concentrations in future studies.

4.4 Six-parameter model of pH and T for the prediction of the chromatographic retention factor

We combined the two variables of temperature and pH into a six-parameter model (as shown in Eq. 6) to explore the combined effect of temperature and pH on chromatographic retention. The fitting parameters a, b, c, d, e, and f (Table 5) of the 57 pentapeptides under the four mobile phase compositions were calculated by an established six-parameter model with pH and T as independent variables and k as the dependent variable to predict the k-value according to different pH and T values. All fitting parameters varied with the change in the mobile phase composition except pH and T. Moreover, a higher proportion of methanol in the mobile phase correlated to a smaller parameter of the acid pentapeptides. Only parameter c displayed an inversely proportional relationship with the proportion of methanol for neutral and basic pentapeptides, and there were no clear trends for other parameters in most cases.

According to the results of the six-parameter model, linear fitting of the experimental k-value and predicted k-value was conducted to assess the prediction capability of chromatographic retention. The R² calculated by linear fitting was used as an evaluation criterion. A random error was present for all data, but the residuals were symmetrically distributed around the axis of y = 0(Supplementary Figure S3). We then fitted the data from five groups of pentapeptides, and the R12 value was just 0.6055, showing the unsatisfactory capacity to predict the chromatographic retention of the studied pentapeptides (Figure 4A). Further classifying the data according to their acid-base properties and fittings, the R₂² value was 0.8603 for the acid pentapeptides (Figure 4B), while both the R_3^2 and R₄² values were lower than 0.7 for the basic and neutral pentapeptides (Figures 4C, D). The results indicated that the sixparameter model had a certain prediction capability for the chromatographic retentions for the acid pentapeptides but was unable to characterize the chromatographic retentions for the basic or neutral pentapeptides. In addition, the R² values fitted by the experimental k-value and predicted k-value under different chromatographic conditions in the six-parameter model of T and pH are shown in Supplementary Table S5. The R² decreased with the increase in column temperature or the methanol volume fraction, indicating that this model was suitable for compounds with higher chromatographic retention.

Internal validation is a commonly used method for evaluating models free of experimental and environmental conditions' limitations (Luo et al., 2020; Vasconcelos et al., 2023). In this study, we used 10-fold cross validation to conduct internal validation. The root mean squared error (RMSE) obtained from 10-fold cross validation was used to evaluate the prediction capability of the models in this study. The average RMSE from the 10 test sets was used to minimize the biased prediction results. The residuals of all pentapeptides and acid pentapeptides were randomly distributed around the y = 0 axis (Supplementary Figure S5). Moreover, the average RMSE of all pentapeptides and acid pentapeptides was 0.48 and 0.20 in the 10 tests (Supplementary Table S7), respectively, indicating that the six-parameter model had both random error and certain prediction capability.

4.5 The six-parameter model of pH and mobile phase compositions for the prediction of the chromatographic retention factor

We considered the combined influence of the mobile phase composition and pH on chromatographic retention by substituting pH and X into Eq. 13 and obtained the fitting parameters A, B, C, D, E, and F of the 57 pentapeptides at three temperatures (Tables 6, 7). X represented φ or P_m^N in the solvent model and was used as the variable to describe the change in the mobile phase. Hence, a sixparameter model was constructed with pH and X as the independent variables and k as the dependent variable, and this model was applied for the prediction of the k-value based on different pH and X values. All fitting parameters varied with the change in temperature expect pH and φ or P_m^N . More importantly, parameter A of acid pentapeptides and parameter C of neutral pentapeptides increased with a decrease in column temperature when φ was the descriptor for the mobile phase composition, while there was no distinct tendency from the other parameters. The six parameters of the 57 pentapeptides deviated from a positive or negative correlation when P_m^N was the descriptor for the mobile phase composition.

The retention factors of the 57 pentapeptides under different elution conditions were predicted with φ or P_m^N as the independent variables according to the six-parameter model, and 4,788 predicted k-values were obtained. We found that all data were distributed regularly around the axis of y = 0 by showing a residual scatter diagram depending on the used independent variable (φ or P_m^N) (Supplementary Figures S4A, B). We further fitted the data from the 57 pentapeptides between the experimental k-values and predicted k-values, and the R² values were used to represent the correlation. When φ was used as the variable to describe the change in the mobile phase, the R_a² value was 0.7367 for all pentapeptides, indicating that the sixparameter model was not extremely suitable for the prediction of the chromatographic retention (Figure 5A). However, the six-parameter model exhibited extraordinary prediction capacity for acid and neutral pentapeptides since the Rb2 value was 0.9718 for acid pentapeptides and the R_d^2 value was 0.9388 for neutral pentapeptides (Figures 5B, D). The Rc2 value was lower than 0.6, indicating that the model was inappropriate for the basic pentapeptides. When P_m^N was used as the variable to describe the change in the mobile phase, the results were very similar to the fitting results of φ (Figure 6A, $R_A^2 = 0.7371$; Figure 6B, $R_B^2 = 0.9722$; Figure 6C, $R_C^2 = 0.7371$; Figure 6D, $R_D^2 = 0.7371$). The correlation between the experimental k-values and predicted k-values was very good, and the model had an excellent capacity to predict the retention of the pentapeptides in RP-HPLC at different mobile phase compositions and pH values, especially for the acid and neutral pentapeptides. Moreover, the model results were always ideal regardless of the used variable (φ or P_m^N) to describe the change in the mobile phase, and the results of P_m^N and pH were slightly better than those of φ and pH, consistent with a previous study. In addition, the prediction ability of the six-parameter model was evaluated under different chromatographic conditions (Supplementary Table S6). Higher T and P_m^N correlated with more relevant results. The correlation of the six-parameter model was inversely proportional to φ. We reached the same conclusion as Section 4.4 specifically, that higher chromatographic retention was beneficial for the prediction capacity of the six-parameter model. Furthermore, 10-fold cross validation was conducted as described in Section 4.4. The residuals of the training and testing sets were randomly distributed around the y = 0 axis (Supplementary Figure S4), and the average RMSE of all pentapeptides was less than 0.8 in the six-parameter model. Moreover, the average RMSE of acid and neutral pentapeptides was approximately 0.3 (Supplementary Table S7), indicating the excellent prediction capacity of the six-parameter model for acid and neutral pentapeptides.

Retention behavior prediction of oligopeptides is valuable for efficient separation and purification. Previous studies have proposed various models to investigate the change in retention based on molecular descriptors or chromatographic theories (Park et al., 2020; Al Musaimi et al., 2023). There are five most commonly used models in studying the effect of mobile phase composition on retention behavior: (1) the linear-solvent-strength model, (2) the quadratic model, (3) the log–log (adsorption) model, (4) the mixed-mode model, and (5) the Neue–Kuss model (den Uijl et al., 2021). These models were able to predict the retention behavior at the first- and second-order levels. However, there was a clear deviation from linearity, especially in the lower organic modifier volume (Baeza-Baeza and García-Alvarez-

Coque, 2020). Furthermore, the QSRR model displayed excellent prediction capacity for ionizable compounds but was limited to the type and calculation method of the molecular descriptors (Kumari et al., 2023). In addition, previous studies have reported the combined influence of temperature and mobile phase composition on chromatographic retention (Arkell et al., 2018; Caltabiano et al., 2018), and the simultaneous effect of pH and temperature or mobile phase composition has been less reported. The six-parameter model of pH and φ or P_m^N performed better in predicting the capacity of acidic compounds than basic compounds in the mixture of water-acetonitrile (Agrafiotou et al., 2011). The combined effect of pH and mobile phase composition also depended on the molecular structure via response surface methodology (D'Archivio and Maggi, 2017). Here, we further proved that the six-parameter model was also suitable for acid or neutral compounds in a water-methanol system regardless of the pH and T or pH and φ or P_m^N were used as independent variables.

5 Conclusion

Herein, we established six-parameter models via RP-HPLC data for predicting the retention factors of pentapeptides under different chromatographic conditions. The relationships of the three parameters pKa, k_{HA} , and k_A for all solutes derived from the sigmoidal model were studied against 1/T, φ , or P_m^N . The results showed that good linear correlations existed between the $logk_{HA}$ and $logk_A$ with respect to either φ or P_m^N for the acid and neutral pentapeptides, and 90% of the R² values were greater than 0.9. Notably, the linear correlations with φ or P_m^N as an independent variable were nearly identical. We then discussed in detail the effect of the distinct elements on the prediction capacity of the models by fitting the experimental k-values and predicted k-values in different groups of pentapeptides or chromatographic conditions. The R² value was 0.8603 and the average RMSE was 0.2 for acid pentapeptides; the R² values were less than 0.7 for the basic and neutral pentapeptides in the sixparameter model of pH and T, indicating that the model was not suitable to predict the change in the chromatographic retention for the basic and neutral pentapeptides at different pH and T. Moreover, the \mathbb{R}^2 values from the models with pH and X as independent variables were greater than 0.93, and the average RMSE was approximately 0.3 for the acid and neutral pentapeptides, indicating an effective prediction capacity of the chromatographic retention.

In this study, there are also some limitations. First, the fitting results of the model would be more reliable with more temperature gradients of chromatographic conditions, but only 3 gradients of column temperature were used in this study. Second, higher chromatographic retention correlated to better fitting results. However, we cannot ensure evident retention results for all studied pentapeptides under other diverse elution conditions. Third, we selected a narrow range of methanol concentrations to produce symmetrical and sharp chromatographic peaks. The methanol concentrations outside this range were undefined as to whether they followed the Soczewiński–Wachtmeister equation. Finally, both six-parameter models showed unsatisfactory prediction capability for the basic pentapeptides, which needs further research. In conclusion, our study determined that the six-parameter model of pH and φ or P_m^N was able to predict the chromatographic retention of the acid and neutral pentapeptides by analyzing the effect of various elements. Furthermore, our study could provide a methodological reference for the analysis and separation of pentapeptides with similar structures and polarities.

Data availability statement

The original contributions presented in the study are included in the article/Supplementary Material; further inquiries can be directed to the corresponding authors.

Author contributions

HP completed the experiment, conducted data processing, and edited the manuscript. XY revised the manuscript and provided valuable input and suggestions. TZ offered important assistance in data processing. HF, ZZ, and JZ supplied the experimental platform and guide. JL and YL critically revised the paper and made amendments and corrections to the manuscript.

Funding

This work is supported by the National Natural Science Foundation of China (Nos. 92057111, 82071538), the Natural Science Foundation of Shaanxi Province (Nos. 2014JM4095, 2018JM7059) and the Xi'an City Science and Technology Project (No. 2017085CG/RC048 (XBDX001)).

References

Abdelhedi, O., and Nasri, M. (2019). Basic and recent advances in marine antihypertensive peptides: Production, structure-activity relationship and bioavailability. *Trends Food Sci. Technol.* 88, 543–557. doi:10.1016/j.tifs.2019. 04.002

Abdelhedi, O., Nasri, R., Jridi, M., Mora, L., Oseguera-Toledo, M. E., Aristoy, M.-C., et al. (2017). *In silico* analysis and antihypertensive effect of ACE-inhibitory peptides from smooth-hound viscera protein hydrolysate: Enzyme-peptide interaction study using molecular docking simulation. *Process Biochem.* 58, 145–159. doi:10.1016/j. procbio.2017.04.032

Agrafiotou, P., Ràfols, C., Castells, C., Bosch, E., and Rosés, M. (2011). Simultaneous effect of pH, temperature and mobile phase composition in the chromatographic retention of ionizable compounds. *J. Chromatogr. A* 1218 (30), 4995–5009. doi:10.1016/j.chroma.2010.12.119

Al Musaimi, O., Valenzo, O. M. M., and Williams, D. R. (2023). Prediction of peptides retention behavior in reversed-phase liquid chromatography based on their hydrophobicity. J. Sep. Sci. 46 (2), 2200743. doi:10.1002/jssc.202200743

Alvarez-Segura, T., Subirats, X., and Rosés, M. (2019). Retention-pH profiles of acids and bases in hydrophilic interaction liquid chromatography. *Anal. Chim. Acta* 1050, 176–184. doi:10.1016/j.aca.2018.11.021

Annadi, A. M., El Zahar, N. M., El-Din, A. A.-S. N., Mohamed, E. H., Mahmoud, S. A., and Attia, M. S. (2022). Development and validation of molnupiravir assessment in bulk powder and pharmaceutical formulation by the RP-HPLC-UV method. *RSC Adv.* 12 (53), 34512–34519. doi:10.1039/D2RA05066H

Arkell, K., Breil, M. P., Frederiksen, S. S., and Nilsson, B. (2018). Mechanistic modeling of reversed-phase chromatography of insulins within the temperature range 10-40 °C. ACS Omega 3 (2), 1946–1954. doi:10.1021/acsomega.7b01527

Attwa, M. W., AlRabiah, H., Mostafa, G. A. E., and Kadi, A. A. (2023). Development of an LC-MS/MS method for quantification of sapitinib in human liver microsomes: *In*

Acknowledgments

The authors greatly appreciate Shaanxi Huikang Biotechnology Co., Ltd., and Xi'an Peihua University for providing the experimental platform.

Conflict of interest

ZZ and JZ were employed by the company Active Protein and Polypeptide Engineering Center of Shaanxi Huikang Biotechnology Co., Ltd. XY was employed by the company Kangya of Ningxia Pharmaceutical Co., Ltd.

The remaining authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

Publisher's note

All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors, and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.

Supplementary material

The Supplementary Material for this article can be found online at: https://www.frontiersin.org/articles/10.3389/fchem.2023.1171824/ full#supplementary-material

silico and in vitro metabolic stability evaluation. Molecules 28 (5), 2322. doi:10.3390/ molecules28052322

Baeza-Baeza, J. J., and García-Alvarez-Coque, M. C. (2020). Extension of the linear solvent strength retention model including a parameter that describes the elution strength changes in liquid chromatography. *J. Chromatogr. A* 1615, 460757. doi:10. 1016/j.chroma.2019.460757

Bergazin, T. D., Tielker, N., Zhang, Y., Mao, J., Gunner, M. R., Francisco, K., et al. (2021). Evaluation of log P, pK(a), and log D predictions from the SAMPL7 blind challenge. J. Comput. Aided. Mol. Des. 35 (7), 771–802. doi:10.1007/s10822-021-00397-3

Besleaga, I., Stepanenko, I., Petrasheuskaya, T. V., Darvasiova, D., Breza, M., Hammerstad, M., et al. (2021). Triapine analogues and their copper(II) complexes: Synthesis, characterization, solution speciation, redox activity, cytotoxicity, and mR2 RNR inhibition. *Inorg. Chem.* 60 (15), 11297–11319. doi:10.1021/acs. inorgchem.1c01275

Biancolillo, A., Maggi, M. A., Bassi, S., Marini, F., and D'Archivio, A. A. (2020). Retention modelling of phenoxy acid herbicides in reversed-phase HPLC under gradient elution. *Molecules* 25 (6), 1262. doi:10.3390/molecules25061262

Caltabiano, A. M., Foley, J. P., and Striegel, A. M. (2018). Organic solvent modifier and temperature effects in non-aqueous size-exclusion chromatography on reversed-phase columns. *J. Chromatogr. A* 1531, 83–103. doi:10.1016/j.chroma. 2017.11.027

D'Archivio, A. A. (2019). Artificial neural network prediction of retention of amino acids in reversed-phase HPLC under application of linear organic modifier gradients and/or pH gradients. *Molecules* 24 (3), 632. doi:10.3390/molecules24030632

D'Archivio, A. A., and Maggi, M. A. (2017). Investigation by response surface methodology of the combined effect of pH and composition of water-methanol mixtures on the stability of curcuminoids. *Food Chem.* 219, 414–418. doi:10.1016/j. foodchem.2016.09.167

den Uijl, M. J., Schoenmakers, P. J., Pirok, B. W. J., and van Bommel, M. R. (2021). Recent applications of retention modelling in liquid chromatography. *J. Sep. Sci.* 44 (1), 88–114. doi:10.1002/jssc.202000905

Elmansi, H., Nasr, J. J., Rageh, A. H., El-Awady, M. I., Hassan, G. S., Abdel-Aziz, H. A., et al. (2019). Assessment of lipophilicity of newly synthesized celecoxib analogues using reversed-phase HPLC. BMC Chem. *[Online]* 13 (1), 84. [Accessed 2019/12//]. doi:10. 1186/s13065-019-0607-6

Faisal, Z., Derdák, D., Lemli, B., Kunsági-Máté, S., Bálint, M., Hetényi, C., et al. (2018). Interaction of 2'R-ochratoxin A with serum albumins: Binding site, effects of site markers, thermodynamics, species differences of albumin-binding, and influence of albumin on its toxicity in mdck cells. *Toxins (Basel)* 10 (9), 353. doi:10.3390/toxins10090353

Fan, K., Peng, J., Peng, H., Zhang, Z., Chen, J., Luo, P., et al. (2022). Effect of spacer alkyl chain length on retention among three imidazolium stationary phases under various modes in high performance liquid chromatography. *J. Chromatogr. A* 1685, 463646. doi:10.1016/j.chroma.2022.463646

Ferreira, M. R., Garzón, A. G., Oliva, M. E., Cian, R. E., Drago, S. R., and D'Alessandro, M. E. (2022). Lipid-lowering effect of microencapsulated peptides from brewer's spent grain in high-sucrose diet-fed rats. *Food Biosci.* 49, 101981. doi:10.1016/j.fbio.2022.101981

Flieger, J., Orzeł, A., Kowalska-Kępczyńska, A., Pizoń, M., Trębacz, H., Majerek, D., et al. (2020). Teicoplanin-Modified HPLC column as a source of experimental parameters for prediction of the anticonvulsant activity of 1,2,4-triazole-3-thiones by the regression models. *Mater. (Basel)* 13 (11), 2650. doi:10.3390/ma13112650

Flieger, J., Trębacz, H., Pizoń, M., Plazińska, A., Plaziński, W., Kowalska, A., et al. (2019). Thermodynamic study of new antiepileptic compounds by combining chromatography on the phosphatidylcholine biomimetic stationary phase and differential scanning calorimetry. J. Sep. Sci. 42 (16), 201900248, doi:10.1002/jssc.201900248

Fouad, M. A., Serag, A., Tolba, E. H., El-Shal, M. A., and El Kerdawy, A. M. (2022). QSRR modeling of the chromatographic retention behavior of some quinolone and sulfonamide antibacterial agents using firefly algorithm coupled to support vector machine. *BMC Chem.* 16 (1), 85. doi:10.1186/s13065-022-00874-2

Galaon, T., and David, V. (2011). Deviation from van't Hoff dependence in RP-LC induced by tautomeric interconversion observed for four compounds. *J. Sep. Sci.* 34 (12), 1423–1428. doi:10.1002/jssc.201100029

Gao, X., Fang, D., Liang, Y., Deng, X., Chen, N., Zeng, M., et al. (2022). Circular RNAs as emerging regulators in COVID-19 pathogenesis and progression. *Front. Immunol.* 13, 980231. doi:10.3389/fimmu.2022.980231

Gisbert-Alonso, A., Navarro-Huerta, J. A., Torres-Lapasió, J. R., and García-Alvarez-Coque, M. C. (2021). Global retention models and their application to the prediction of chromatographic fingerprints. *J. Chromatogr. A* 1637, 461845. doi:10.1016/j.chroma. 2020.461845

Guo, H., Li, L., and Gao, L. (2023). Paraquat and diquat: Recent updates on their pretreatment and analysis methods since 2010 in biological samples. *Molecules* 28 (2), 684. doi:10.3390/molecules28020684

Guo, H., Wahab, M. F., Berthod, A., and Armstrong, D. W. (2018). Mass spectrometry detection of basic drugs in fast chiral analyses with vancomycin stationary phases. *J. Pharm. Anal.* 8 (5), 324–332. doi:10.1016/j.jpha.2018.08.001

Hong, X., Zhao, Y., Zhuang, R., Liu, J., Guo, G., Chen, J., et al. (2020). Bioremediation of tetracycline antibiotics-contaminated soil by bioaugmentation. *RSC Adv.* 10 (55), 33086–33102. doi:10.1039/D0RA04705H

Huang, W. W., Hong, B. H., Sun, J. P., Tan, R., Bai, K. K., Yang, T., et al. (2019). Comparing the simultaneous determination of cis- and trans-palmitoleic acid in fish oil using HPLC and GC. *Lipids Health Dis.* 18 (1), 86. doi:10.1186/s12944-019-1033-4

Idroes, R., Muslem, M. S., Idroes, G. M., Suhendra, R., Suhendra, R., et al. (2020). The effect of column and temperature variation on the determination of the dead time in gas chromatographic systems using indirect methods. *Heliyon* 6 (2), e03302. doi:10.1016/j. heliyon.2020.e03302

Janicka, M., Sztanke, M., and Sztanke, K. (2020). Predicting the blood-brain barrier permeability of new drug-like compounds via HPLC with various stationary phases. *Molecules* 25 (3), 487. doi:10.3390/molecules25030487

Kaczmarski, K., and Chutkowski, M. (2021). Impact of changes in physicochemical parameters of the mobile phase along the column on the retention time in gradient liquid chromatography. Part A – temperature gradient. *J. Chromatogr. A* 1655, 462509. doi:10.1016/j.chroma.2021.462509

Konçe, İ., Demiralay, E. Ç., and Ortak, H. Y. (2019). Chromatographic determination of thermodynamic acid dissociation constants of tetracycline antibiotics and their epimers. *J. Chromatogr. Sci.* 57 (8), 745–750. doi:10.1093/chromsci/bmz051

Kumari, P., Van Laethem, T., Hubert, P., Fillet, M., Sacré, P. Y., and Hubert, C. (2023). Quantitative structure retention-relationship modeling: Towards an innovative general-purpose strategy. *Molecules* 28 (4), 1696. doi:10.3390/molecules28041696

Langyan, S., Khan, F. N., Yadava, P., Alhazmi, A., Mahmoud, S. F., Saleh, D. I., et al. (2021). *In silico* proteolysis and analysis of bioactive peptides from sequences of fatty acid desaturase 3 (FAD3) of flaxseed protein. *Saudi. J. Biol. Sci.* 28 (10), 5480–5489. doi:10.1016/j.sjbs.2021.08.027

Lin, T., Chen, B., Fang, L., You, H., Chu, C., Shao, Q., et al. (2022). Solvent strength of organic phase for two biphasic solvent systems in high speed countercurrent chromatography. *J. Chromatogr. A* 1680, 463422. doi:10.1016/j. chroma.2022.463422

Liu, C., Wang, W., Zhang, K., Liu, Q., Ma, T., Tan, L., et al. (2022). Protective effects of polydatin from grapes and reynoutria japonica houtt. On damaged macrophages treated with acetaminophen. *Nutrients* 14 (10), 2077. doi:10.3390/nu14102077

Luo, Y., Chalkou, K., Yamada, R., Funada, S., Salanti, G., and Furukawa, T. A. (2020). Predicting the treatment response of certolizumab for individual adult patients with rheumatoid arthritis: Protocol for an individual participant data meta-analysis. *Syst. Rev.* 9 (1), 140. doi:10.1186/s13643-020-01401-x

Marchetti, N., Giovannini, P. P., Catani, M., Pasti, L., and Cavazzini, A. (2019). Thermodynamic insights into the separation of carotenoids in reversed-phase liquid chromatography. *Int. J. Anal. Chem.* 2019, 1. 7. doi:10.1155/2019/7535813

Nagase, K., Umemoto, Y., and Kanazawa, H. (2021). Effect of pore diameter on the elution behavior of analytes from thermoresponsive polymer grafted beads packed columns. *Sci. Rep.* 11 (1), 9976. doi:10.1038/s41598-021-89165-9

Nie, Y., Li, J., Yang, X., Hou, X., and Fang, H. (2022). Development of QSRR model for hydroxamic acids using PCA-GA-BP algorithm incorporated with molecular interaction-based features. *Front. Chem.* 10, 1056701. doi:10.3389/fchem.2022. 1056701

Numviyimana, C., Chmiel, T., Kot-Wasik, A., and Namieśnik, J. (2019). Study of pH and temperature effect on lipophilicity of catechol-containing antioxidants by reversed phase liquid chromatography. *Microchem. J.* 145, 380–387. doi:10.1016/j. microc.2018.10.048

Ofosu, F. K., Mensah, D. F., Daliri, E. B., and Oh, D. H. (2021). Exploring molecular insights of cereal peptidic antioxidants in metabolic syndrome prevention. *Antioxidants (Basel)* 10 (4), 518. doi:10.3390/antiox10040518

Oney-Montalvo, J. E., Morozova, K., Ramírez-Sucre, M. O., Scampicchio, M., and Rodríguez-Buenfil, I. M. (2022). Determination of peak purity in HPLC by coupling coulometric array detection and two-dimensional correlation analysis. *Sensors (Basel)* 22 (5), 1794. doi:10.3390/s22051794

Park, S. H., De Pra, M., Haddad, P. R., Grosse, S., Pohl, C. A., and Steiner, F. (2020). Localised quantitative structure-retention relationship modelling for rapid method development in reversed-phase high performance liquid chromatography. *J. Chromatogr. A* 1609, 460508. doi:10.1016/j.chroma.2019.460508

Phyo, Y. Z., Cravo, S., Palmeira, A., Tiritan, M. E., Kijjoa, A., Pinto, M. M. M., et al. (2018). Enantiomeric resolution and docking studies of chiral xanthonic derivatives on chirobiotic columns. *Molecules* 23 (1), 142. doi:10.3390/ molecules23010142

Qiao, Q. Q., Luo, Q. B., Suo, S. K., Zhao, Y. Q., Chi, C. F., and Wang, B. (2022). Preparation, characterization, and cytoprotective effects on HUVECs of fourteen novel angiotensin-I-converting enzyme inhibitory peptides from protein hydrolysate of tuna processing by-products. *Front. Nutr.* 9, 868681. doi:10.3389/ fnut.2022.868681

Rook, M. L., Musgaard, M., and MacLean, D. M. (2021). Coupling structure with function in acid-sensing ion channels: Challenges in pursuit of proton sensors. *J. Physiol.* 599 (2), 417–430. doi:10.1113/JP278707

Samtiya, M., Acharya, S., Pandey, K. K., Aluko, R. E., Udenigwe, C. C., and Dhewa, T. (2021). Production, purification, and potential health applications of edible seeds' bioactive peptides: A concise review. *Foods* 10 (11), 2696. doi:10.3390/foods10112696

Sheng, Y., Qiu, Y. T., Wang, Y. M., Chi, C. F., and Wang, B. (2022). Novel antioxidant collagen peptides of siberian sturgeon (acipenserbaerii) cartilages: The preparation, characterization, and cytoprotection of H(2)O(2)-damaged human umbilical vein endothelial cells (HUVECs). *Mar. Drugs* 20 (5), 325. doi:10.3390/md20050325

Shi, Y., Hu, J., Wang, H., Yan, Z., Zhao, G., Gao, X., et al. (2022). Establishing a UHPLC-MS/MS method for evaluation of the influence of stir-frying on the pharmacokinetics of seven compounds in Arctii Fructus. *RSC Adv.* 12 (42), 27525–27533. doi:10.1039/D2RA03637A

Sitkov, N., Zimina, T., Kolobov, A., Sevostyanov, E., Trushlyakova, V., Luchinin, V., et al. (2021). Study of the fabrication Technology of hybrid microfluidic biochips for label-free detection of proteins. *Micromachines (Basel)* 13 (1), 20. doi:10.3390/ mil3010020

Soriano-Meseguer, S., Fuguet, E., Port, A., and Rosés, M. (2019). Influence of the acidbase ionization of drugs in their retention in reversed-phase liquid chromatography. *Anal. Chim. Acta* 1078, 200–211. doi:10.1016/j.aca.2019.05.063

Sousa, H. B. A., Martins, C. S. M., and Prior, J. A. V. (2021). You don't learn that in school: An updated practical guide to carbon quantum dots. *Nanomater. (Basel)* 11 (3), 611. doi:10.3390/nano11030611

Suo, S. K., Zhao, Y. Q., Wang, Y. M., Pan, X. Y., Chi, C. F., and Wang, B. (2022). Seventeen novel angiotensin converting enzyme (ACE) inhibitory peptides from the protein hydrolysate of *Mytilus edulis*: Isolation, identification, molecular docking study, and protective function on HUVECs. *Food Funct.* 13 (14), 7831–7846. doi:10.1039/ d2fo00275b Tanase, M., Soare, A., David, V., and Moldoveanu, S. C. (2019). Sources of nonlinear van't Hoff temperature dependence in high-performance liquid chromatography. *ACS Omega* 4 (22), 19808–19817. doi:10.1021/acsomega.9b02689

Tengattini, S., Rimaroli, C., Galmozzi, M. R., Furlanetto, S., Massolini, G., and Temporini, C. (2022). Effect of mobile phase pH on liquid chromatography retention of mepartricin related compounds and impurities as support to the structural investigation by liquid chromatography-mass spectrometry. *J. Pharm. Biomed.* 220, 114971. doi:10.1016/j.jpba.2022.114971

Tsui, H.-W., Kuo, C.-H., and Huang, Y.-C. (2019). Elucidation of retention behaviors in reversed-phase liquid chromatography as a function of mobile phase composition. *J. Chromatogr. A* 1595, 127–135. doi:10.1016/j.chroma.2019.02.049

Vasconcelos, L., Dias, L. G., Leite, A., Ferreira, I., Pereira, E., Silva, S., et al. (2023). SVM regression to assess meat characteristics of bísaro pig loins using NIRS methodology. *Foods* 12 (3), 470. doi:10.3390/foods12030470

Waili, Y., Gahafu, Y., Aobulitalifu, A., Chang, Z., Xie, X., and Kawuli, G. (2021). Isolation, purification, and characterization of antioxidant peptides from fresh mare's milk. *Food Sci. Nutr.* 9 (7), 4018–4027. doi:10.1002/fsn3.2292

Wang, J., Wu, Y., Chen, Z., Chen, Y., Lin, Q., and Liang, Y. (2022). Exogenous bioactive peptides have a potential therapeutic role in delaying aging in rodent models. *Int. J. Mol. Sci.* 23 (3), 1421. doi:10.3390/ijms23031421

Xie, Y., Yuan, P., Heng, T., Du, L., An, Q., Zhang, B., et al. (2022). Insight into the formation of cocrystal and salt of tenoxicam from the isomer and conformation. *Pharmaceutics* 14 (9), 1968. doi:10.3390/pharmaceutics14091968

Xiong, J., Li, Z., Wang, G., Fu, Z., Zhong, F., Xu, T., et al. (2022). Multi-instance learning of graph neural networks for aqueous pKa prediction. *Bioinformatics* 38 (3), 792–798. doi:10.1093/bioinformatics/btab714

Xu, Z., Chughtai, H., Tian, L., Liu, L., Roy, J. F., and Bayen, S. (2023). Development of quantitative structure-retention relationship models to improve the identification of leachables in food packaging using non-targeted analysis. *Talanta* 253, 123861. doi:10. 1016/j.talanta.2022.123861

Yang, J., Moraga, A., Xu, J., Zhao, Y., Luo, P., Lao, K. H., et al. (2020a). A histone deacetylase 7-derived peptide promotes vascular regeneration via facilitating 14-3-3 γ phosphorylation. *Stem Cells* 38 (4), 556–573. doi:10.1002/stem.3122

Yang, X., Peng, H., Han, N., Zhang, Z., Bai, X., Zhao, T., et al. (2020b). Quantitative structure-chromatographic retention relationship of synthesized peptides (HGRFG,

NPNPT) and their derivatives. Anal. Biochem. 597, 113653. doi:10.1016/j.ab.2020. 113653

Yang, Y.-X., Zhang, Q., Li, Q.-Q., Xia, Z.-N., Chen, H., Zhou, K., et al. (2018). pHdependent surface electrostatic effects in retention on immobilized artificial membrane chromatography: Determination of the intrinsic phospholipid-water sorption coefficients of diverse analytes. *J. Chromatogr. A* 1570, 172–182. doi:10.1016/j.chroma.2018.07.081

Ye, H., Xu, Y., Sun, Y., Liu, B., Chen, B., Liu, G., et al. (2023). Purification, identification and hypolipidemic activities of three novel hypolipidemic peptides from tea protein. *Food Res. Int.* 165, 112450. doi:10.1016/j.foodres.2022.112450

Yılmaz Ortak, H., and Cubuk Demiralay, E. (2019). Effect of temperature on the retention of Janus kinase 3 inhibitor in different mobile phase compositions using reversed-phase liquid chromatography. *J. Pharm. Biomed. Anal.* 164, 706–712. doi:10. 1016/j.jpba.2018.11.032

Yuan, N., Chen, J., Cai, T., Li, Z., Guan, M., Zhao, L., et al. (2020). Glucose-based carbon dots-modified silica stationary phase for hydrophilic interaction chromatography. *J. Chromatogr. A* 1619, 460930. doi:10.1016/j.chroma.2020.460930

Żesławska, E., Zakrzewski, R., Nowicki, A., Korona-Głowniak, I., Lyčka, A., Kania, A., et al. (2022). Synthesis, crystal structures, lipophilic properties and antimicrobial activity of 5-Pyridylmethylidene-3-rhodanine-carboxyalkyl acids derivatives. *Molecules* 27 (13), 3975. doi:10.3390/molecules27133975

Zhang, Y., Liang, R., Xie, A., Shi, W., Huang, H., and Zhong, Y. (2020). Antagonistic peptides that specifically bind to the first and second extracellular loops of CCR5 and anti-IL-23p19 antibody reduce airway inflammation by suppressing the IL-23/ Th17 signaling pathway. *Mediat. Inflamm.* 2020, 1–13. doi:10.1155/2020/1719467

Zhang, Y., Zhao, Z., Wang, K., Lyu, K., Yao, C., Li, L., et al. (2022). Molecular docking assisted exploration on solubilization of poorly soluble drug remdesivir in sulfobutyl ether-tycyclodextrin. *AAPS Open* 8 (1), 9. doi:10.1186/s41120-022-00054-5

Zhu, X., Li, P., Tang, J., Su, Y., Xiao, M., Xue, H., et al. (2022). A simple and practical solvent system selection strategy for high-speed countercurrent chromatography based on the HPLC polarity parameter model. *Anal. Methods* 14 (46), 4822–4831. doi:10. 1039/D2AY01377K

Yilmaz, H. (2021). Janus Kinase İnhibitörünün Stabilitesini Değerlendirmeye Yönelik RPLC Metodu Kullanılarak Yapılan Zorunlu Bozunma Çalışmaları. *SDÜ Fen. Bil. Enst. Der.*, 22, 134-141. doi:10.19113/sdufenbed.826534